

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032919\
 Data File : PD052308.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Mar 2019 15:56
 Operator : AJ\SJ
 Sample : PB118193BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 04:56:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 29 13:06:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.222	4.043	23551464	524.0E6	16.878	18.078
28)	SA Decachlor...	7.819	9.156	25369219	305.5E6	19.035	19.643
Target Compounds							
3)	MA gamma-BHC...	3.919	0.000	3783488	0	1.730	N.D. #
8)	B Heptachlo...	0.000	5.918	0	7264876	N.D.	0.225 #
10)	B gamma-Chl...	0.000	6.187f	0	9601663	N.D.	0.300 #
11)	B alpha-Chl...	0.000	6.233	0	1753258	N.D.	0.056 #
15)	B Endosulfa...	0.000	6.982	0	12797277	N.D.	0.514 #
17)	MA 4,4'-DDT	0.000	7.207f	0	2286059	N.D.	0.123 #
21)	B Endrin ke...	0.000	7.814f	0	2071793	N.D.	0.101 #
25)	Chlordane-3	0.000	6.187f	0	9601663	N.D.	2.176 #
26)	Chlordane-4	0.000	6.233	0	1753258	N.D.	0.345 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032919\
Data File : PD052308.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Mar 2019 15:56
Operator : AJ\SJ
Sample : PB118193BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 04:56:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
Quant Title : GC Extractables
QLast Update : Fri Mar 29 13:06:45 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.222 min
Delta R.T.: 0.000 min
Response: 23551464
Conc: 16.88 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 523964113
Conc: 18.08 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 3.919 min
Delta R.T.: 0.009 min
Response: 3783488
Conc: 1.73 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 0.000 min
Exp R.T. : 4.723 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 4.846 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.918 min
Delta R.T.: -0.012 min
Response: 7264876
Conc: 0.23 ng/ml




#10 gamma-Chlordane

R.T.: 0.000 min
Exp R.T.: 5.063 min
Response: 0
Conc: N.D.



#10 gamma-Chlordane


R.T.: 6.187 min
Delta R.T.: 0.022 min
Response: 9601663
Conc: 0.30 ng/ml



#11 alpha-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.119 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#11 alpha-Chlordane

R.T.: 6.233 min
Delta R.T.: -0.005 min
Response: 1753258
Conc: 0.06 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 5.919 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.982 min
Delta R.T.: 0.006 min
Response: 12797277
Conc: 0.51 ng/ml



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T.: 6.021 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.207 min
Delta R.T.: 0.019 min
Response: 2286059
Conc: 0.12 ng/ml




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T.: 6.770 min
Response: 0
Conc: N.D.



#21 Endrin ketone


R.T.: 7.814 min
Delta R.T.: 0.015 min
Response: 2071793
Conc: 0.10 ng/ml



#25 Chlordane-3


R.T.: 0.000 min
Exp R.T.: 5.064 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#25 Chlordane-3

R.T.: 6.187 min
Delta R.T.: 0.021 min
Response: 9601663
Conc: 2.18 ng/ml




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T.: 5.118 min
Response: 0
Conc: N.D.



#26 Chlordane-4


R.T.: 6.233 min
Delta R.T.: -0.007 min
Response: 1753258
Conc: 0.34 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.819 min
Delta R.T.: 0.000 min
Response: 25369219
Conc: 19.04 ng/ml

Instrument :
ECD_D
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.156 min
Delta R.T.: 0.000 min
Response: 305478514
Conc: 19.64 ng/ml