

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032919\
 Data File : PD052323.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Mar 2019 19:27
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 3/30/2019 10:13:30 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 04:58:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 29 13:06:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.221	4.042	65245587	1314.8E6	46.757	45.363
28)	SA Decachlor...	7.818	9.155	64223104	770.1E6	48.188	49.522
Target Compounds							
2)	A alpha-BHC	3.641	4.435	109.2E6	2047.9E6	48.018	45.499
3)	MA gamma-BHC...	3.909	4.722	103.9E6	1842.6E6	47.497	45.803
4)	MA Heptachlor	4.186	5.235	103.2E6	1814.6E6	48.592	45.736
5)	MB Aldrin	4.417	5.542	97575132	1700.5E6	48.460	45.925
6)	B beta-BHC	4.155	4.892	41494029	822.0E6	47.013	45.430
7)	B delta-BHC	4.343	5.111	98339979	1685.1E6	48.214	51.211
8)	B Heptachlo...	4.845	5.930	91059821	1573.2E6	47.913	48.738
9)	A Endosulfan I	5.166	6.289	85002753	1378.3E6	47.564	47.181
10)	B gamma-Chl...	5.062	6.164	93117470	1528.3E6	47.755	47.768m
11)	B alpha-Chl...	5.118	6.238	92131609	1497.6E6	47.735	47.712
12)	B 4,4'-DDE	5.288	6.394	87502797	1423.1E6	46.924	46.616
13)	MA Dieldrin	5.402	6.547	93730622	1434.1E6	48.476	47.945
14)	MA Endrin	5.649	6.766	83830066	1141.2E6	45.527	45.543
15)	B Endosulfa...	5.918	6.975	81599569	1212.6E6	47.540	48.664
16)	A 4,4'-DDD	5.789	6.886	71412550	1160.3E6	46.723	48.514
17)	MA 4,4'-DDT	6.019	7.186	70834402	1001.0E6	49.239	53.720
18)	B Endrin al...	6.085	7.100	68255077	1014.8E6	47.610	49.296
19)	B Endosulfa...	6.293	7.324	72671181	1091.7E6	47.391	49.624
20)	A Methoxychlor	6.565	7.646	40820272	524.7E6	48.318	52.457
21)	B Endrin ke...	6.768	7.798	89958168	1077.7E6	49.461	52.597
22)	Mirex	6.937	8.255	61199941	677.2E6	49.957	52.518

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032919\
Data File : PD052323.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Mar 2019 19:27
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
3/30/2019 10:13:30 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 04:58:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
Quant Title : GC Extractables
QLast Update : Fri Mar 29 13:06:45 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

