

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032919\
 Data File : PD052294.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Mar 2019 12:40
 Operator : AJ\SJ
 Sample : PCHLORICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 29 13:03:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 29 13:01:48 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.221	4.043	7004751	167.4E6	4.765	5.841
28)	SA Decachlor...	7.818	9.157	7931569	86549573	5.812	5.551
Target Compounds							
23)	Chlordane-1	4.043	5.045	2906359	86882932	45.750	58.329 #
24)	Chlordane-2	4.523	5.515	3990340	82336806	54.136	61.083
25)	Chlordane-3	5.063	6.165	10039246	251.2E6	44.949	58.851 #
26)	Chlordane-4	5.118	6.239	10022876	295.1E6	46.828	60.289 #
27)	Chlordane-5	5.922	7.050	2729012	46896166	40.360	50.118



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032919\
Data File : PD052294.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Mar 2019 12:40
Operator : AJ\SJ
Sample : PCHLORICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 29 13:03:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
Quant Title : GC Extractables
QLast Update : Fri Mar 29 13:01:48 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.221 min
Delta R.T.: 0.000 min
Response: 7004751
Conc: 4.77 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050




#1 Tetrachloro-m-xylene

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 167413349
Conc: 5.84 ng/ml




#23 Chlordane-1

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 2906359
Conc: 45.75 ng/ml



#23 Chlordane-1


R.T.: 5.045 min
Delta R.T.: 0.000 min
Response: 86882932
Conc: 58.33 ng/ml



#24 Chlordane-2


R.T.: 4.523 min
Delta R.T.: 0.000 min
Response: 3990340
Conc: 54.14 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050




#24 Chlordane-2

R.T.: 5.515 min
Delta R.T.: 0.000 min
Response: 82336806
Conc: 61.08 ng/ml




#25 Chlordane-3

R.T.: 5.063 min
Delta R.T.: -0.001 min
Response: 10039246
Conc: 44.95 ng/ml



#25 Chlordane-3


R.T.: 6.165 min
Delta R.T.: 0.000 min
Response: 251154883
Conc: 58.85 ng/ml



#26 Chlordane-4


R.T.: 5.118 min
Delta R.T.: 0.000 min
Response: 10022876
Conc: 46.83 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050




#26 Chlordane-4

R.T.: 6.239 min
Delta R.T.: -0.001 min
Response: 295092003
Conc: 60.29 ng/ml




#27 Chlordane-5

R.T.: 5.922 min
Delta R.T.: -0.001 min
Response: 2729012
Conc: 40.36 ng/ml



#27 Chlordane-5


R.T.: 7.050 min
Delta R.T.: 0.000 min
Response: 46896166
Conc: 50.12 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.818 min
Delta R.T.: 0.000 min
Response: 7931569
Conc: 5.81 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050



#28 Decachlorobiphenyl

R.T.: 9.157 min
Delta R.T.: -0.001 min
Response: 86549573
Conc: 5.55 ng/ml