

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032919\
 Data File : PD052316.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Mar 2019 17:48
 Operator : AJ\SJ
 Sample : K1825-11DL 5X
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PT-CHLR-WPDL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 04:58:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 29 13:06:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.220	4.043	7062412	135.4E6	5.061	4.671
28)	SA Decachlor...	7.818	9.157	7405649	86662199	5.557	5.573
Target Compounds							
23)	Chlordane-1	4.043	5.046	28320018	657.2E6	450.999	428.365
24)	Chlordane-2	4.522	5.516	33680728	627.3E6	452.599	447.943
25)	Chlordane-3	5.063	6.166	100.1E6	1991.3E6	452.661	451.245
26)	Chlordane-4	5.119	6.240	96008681	2279.6E6	452.673	448.291
27)	Chlordane-5	5.923	7.050	30429448	455.6E6	467.907	493.870



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032919\
Data File : PD052316.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Mar 2019 17:48
Operator : AJ\SJ
Sample : K1825-11DL 5X
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PT-CHLR-WPDL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 04:58:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
Quant Title : GC Extractables
QLast Update : Fri Mar 29 13:06:45 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.220 min
Delta R.T.: 0.000 min
Response: 7062412
Conc: 5.06 ng/ml

Instrument :
ECD_D
ClientSampleId :
PT-CHLR-WPDL




#1 Tetrachloro-m-xylene

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 135391916
Conc: 4.67 ng/ml




#23 Chlordane-1

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 28320018
Conc: 451.00 ng/ml



#23 Chlordane-1


R.T.: 5.046 min
Delta R.T.: 0.000 min
Response: 657154675
Conc: 428.37 ng/ml



#24 Chlordane-2


R.T.: 4.522 min
Delta R.T.: 0.000 min
Response: 33680728
Conc: 452.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
PT-CHLR-WPDL




#24 Chlordane-2

R.T.: 5.516 min
Delta R.T.: 0.000 min
Response: 627330398
Conc: 447.94 ng/ml




#25 Chlordane-3

R.T.: 5.063 min
Delta R.T.: 0.000 min
Response: 100100256
Conc: 452.66 ng/ml



#25 Chlordane-3


R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 1991251551
Conc: 451.25 ng/ml



#26 Chlordane-4


R.T.: 5.119 min
Delta R.T.: 0.000 min
Response: 96008681
Conc: 452.67 ng/ml

Instrument :
ECD_D
ClientSampleId :
PT-CHLR-WPDL




#26 Chlordane-4

R.T.: 6.240 min
Delta R.T.: 0.000 min
Response: 2279621301
Conc: 448.29 ng/ml




#27 Chlordane-5

R.T.: 5.923 min
Delta R.T.: 0.000 min
Response: 30429448
Conc: 467.91 ng/ml



#27 Chlordane-5


R.T.: 7.050 min
Delta R.T.: 0.000 min
Response: 455649117
Conc: 493.87 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.818 min
Delta R.T.: 0.000 min
Response: 7405649
Conc: 5.56 ng/ml

Instrument :
ECD_D
ClientSampleId :
PT-CHLR-WPDL



#28 Decachlorobiphenyl

R.T.: 9.157 min
Delta R.T.: 0.000 min
Response: 86662199
Conc: 5.57 ng/ml