

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032921\
 Data File : PD061834.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Mar 2021 21:32
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICPD032921

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 05:37:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX032921.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 30 05:22:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	793.770	741.429 E3	6.6	94	0.00
2	Toxaphene-1	10.655	10.503 E3	1.4	91	0.00
3	Toxaphene-2	8.803	8.556 E3	2.8	95	0.00
4	Toxaphene-3	19.331	18.066 E3	6.5	92	0.00
5	Toxaphene-4	6.930	6.393 E3	7.7	90	0.00
6	Toxaphene-5	8.628	7.532 E3	12.7	95	0.00
7 SA	Decachlorobiphenyl	1.194	1.066 E6	10.7	92	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	8.205	7.669 E6	6.5	94	0.00
2	Toxaphene-1	105.244	96.512 E3	8.3	95	0.00
3	Toxaphene-2	87.842	83.267 E3	5.2	95	0.00
4	Toxaphene-3	248.836	233.955 E3	6.0	93	0.00
5	Toxaphene-4	157.363	144.963 E3	7.9	94	0.00
6	Toxaphene-5	108.747	97.766 E3	10.1	88	0.00
7 SA	Decachlorobiphenyl	8.981	8.188 E6	8.8	93	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

(#) = Out of Range

SPCC's out = 0 CCC's out = 0