

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD033121\
 Data File : PD061916.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Mar 2021 10:09
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 4/1/2021 1:47:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 31 14:25:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032921.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 30 04:47:56 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.486	3.981	44384423	482.3E6	53.009	55.510
28)	SA Decachlor...	8.258	9.040	55116543	395.3E6	44.795	41.870
Target Compounds							
2)	A alpha-BHC	3.925	4.370	68749863	774.4E6	54.433	57.571
3)	MA gamma-BHC...	4.210	4.654	65138601	708.3E6	54.864	56.786
4)	MA Heptachlor	4.513	5.163	65541520	719.9E6	56.299	55.242
5)	MB Aldrin	4.764	5.468	63437750	674.8E6	52.859	54.222
6)	B beta-BHC	4.460	4.823	28669465	294.2E6	52.574	56.957
7)	B delta-BHC	4.667	5.040	66003551	641.3E6	54.332	48.485
8)	B Heptachlo...	5.208	5.853	59631924	591.0E6	52.019	49.908
9)	A Endosulfan I	5.549	6.210	56333238	547.9E6	51.300	50.540
10)	B gamma-Chl...	5.435	6.087	60702772	605.0E6	51.823	51.032
11)	B alpha-Chl...	5.494	6.160	59680483	576.6E6	51.359	50.576
12)	B 4,4'-DDE	5.657	6.316	59342898	574.7E6	52.812	49.867
13)	MA Dieldrin	5.793	6.466	59403236	569.9E6	51.392	49.376
14)	MA Endrin	6.051	6.684	54466207	486.9E6	53.008	49.204
15)	B Endosulfa...	6.325	6.893	50661378	441.8E6	49.431	46.535
16)	A 4,4'-DDD	6.175	6.805	49143450	433.4E6	52.807	45.770m
17)	MA 4,4'-DDT	6.417	7.105	49310813	470.8E6	54.746	49.290
18)	B Endrin al...	6.493	7.017	41371849	361.3E6	48.424	45.944
19)	B Endosulfa...	6.707	7.240	52716059	457.3E6	48.855	45.631
20)	A Methoxychlor	6.963	7.563	28742512	218.0E6	60.831	48.358
21)	B Endrin ke...	7.200	7.710	60911897	436.5E6	47.727	43.041
22)	Mirex	7.391	8.165	43419235	277.1E6	52.032	42.997

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD033121\
Data File : PD061916.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Mar 2021 10:09
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
4/1/2021 1:47:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 14:25:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032921.M
Quant Title : GC Extractables
QLast Update : Tue Mar 30 04:47:56 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

