

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040119\
 Data File : PD052412.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Apr 2019 9:49
 Operator : AJ\SJ
 Sample : PTOXCCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXCCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 11:37:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX032919.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 29 14:07:32 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.221	4.043	63183089	1363.5E6	46.504	48.877
7) SA Decachlor...	7.818	9.158	59336509	783.6E6	44.179	51.155
Target Compounds						
2) Toxaphene-1	5.371	6.446	3882423	73986000	470.838	533.320
3) Toxaphene-2	5.951	6.738	13421302	91544922	489.735	559.579
4) Toxaphene-3	6.191	7.224	13396262	324.5E6	515.407	561.978
5) Toxaphene-4	6.536	7.311	9423657	196.9E6	493.435	576.385
6) Toxaphene-5	6.664	7.727	16017766	137.9E6	488.051	562.772


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040119\
Data File : PD052412.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Apr 2019 9:49
Operator : AJ\SJ
Sample : PTOXCCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 11:37:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX032919.M
Quant Title : GC Extractables
QLast Update : Fri Mar 29 14:07:32 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.221 min
Delta R.T.: -0.001 min
Response: 63183089
Conc: 46.50 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#1 Tetrachloro-m-xylene

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 1363539422
Conc: 48.88 ng/ml




#2 Toxaphene-1

R.T.: 5.371 min
Delta R.T.: 0.000 min
Response: 3882423
Conc: 470.84 ng/ml



#2 Toxaphene-1


R.T.: 6.446 min
Delta R.T.: 0.000 min
Response: 73986000
Conc: 533.32 ng/ml



#3 Toxaphene-2


R.T.: 5.951 min
Delta R.T.: -0.001 min
Response: 13421302
Conc: 489.74 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#3 Toxaphene-2

R.T.: 6.738 min
Delta R.T.: 0.000 min
Response: 91544922
Conc: 559.58 ng/ml




#4 Toxaphene-3

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 13396262
Conc: 515.41 ng/ml



#4 Toxaphene-3


R.T.: 7.224 min
Delta R.T.: 0.000 min
Response: 324499450
Conc: 561.98 ng/ml



#5 Toxaphene-4


R.T.: 6.536 min
Delta R.T.: -0.001 min
Response: 9423657
Conc: 493.43 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#5 Toxaphene-4

R.T.: 7.311 min
Delta R.T.: 0.000 min
Response: 196859787
Conc: 576.39 ng/ml




#6 Toxaphene-5

R.T.: 6.664 min
Delta R.T.: 0.000 min
Response: 16017766
Conc: 488.05 ng/ml



#6 Toxaphene-5


R.T.: 7.727 min
Delta R.T.: 0.000 min
Response: 137929507
Conc: 562.77 ng/ml



#7 Decachlorobiphenyl

R.T.: 7.818 min
Delta R.T.: -0.001 min
Response: 59336509
Conc: 44.18 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500



#7 Decachlorobiphenyl

R.T.: 9.158 min
Delta R.T.: 0.001 min
Response: 783645827
Conc: 51.15 ng/ml