

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD040219CLP.M
 Title : GC Extractables
 Last Update : Mon Apr 01 19:05:44 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD052439.D 1 =PD052435.D 2 =PD052437.D
 4 =PD052441.D 5 =PD052443.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.372	1.350	1.364	1.407	1.582	1.415 E6	6.76
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin	2.031	1.906	1.946	2.162	2.473	2.104 E6	10.87
6) B	beta-BHC	0.951	1.001	0.987	0.925	1.026	0.978 E6	4.13
7) B	delta-BHC	2.137	2.004	2.051	2.275	2.550	2.203 E6	9.96
8) B	Heptachlor epoxide	1.959	1.924	1.920	2.035	2.227	2.013 E6	6.38
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane	1.994	1.937	1.940	2.079	2.339	2.058 E6	8.14
11) B	cis-Chlordane	1.951	1.927	1.923	2.019	2.218	2.008 E6	6.15
12) B	4,4'-DDE	1.946	1.730	1.825	2.100	2.433	2.007 E6	13.73
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II	1.845	1.785	1.796	1.884	2.054	1.873 E6	5.83
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde	1.468	1.479	1.460	1.500	1.584	1.498 E6	3.35
19) B	Endosulfan Sulfate	1.651	1.696	1.659	1.717	1.759	1.696 E6	2.60
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone	1.898	1.833	1.863	1.933	2.039	1.913 E6	4.16
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.454	1.692	1.546	1.391	1.452	1.507 E6	7.77

Signal #2 Calibration Files

3 =PD052439.D 1 =PD052435.D 2 =PD052437.D
 4 =PD052441.D 5 =PD052443.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.104	3.426	3.329	2.904	3.191	3.191 E7	6.35
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin	4.096	4.545	4.370	3.858	4.064	4.187 E7	6.46
6) B	beta-BHC	2.018	2.301	2.226	1.865	1.960	2.074 E7	8.84
7) B	delta-BHC	4.257	4.515	4.619	4.045	4.486	4.385 E7	5.27
8) B	Heptachlor epoxide	3.764	4.262	4.055	3.502	3.475	3.812 E7	9.03
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane	3.750	4.123	3.993	3.523	3.515	3.781 E7	7.24
11) B	cis-Chlordane	3.627	4.020	3.817	3.280	3.319	3.613 E7	8.80
12) B	4,4'-DDE	3.407	3.663	3.568	3.225	3.197	3.412 E7	6.02
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II	2.854	3.162	3.020	2.657	2.547	2.848 E7	8.87
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde	2.362	2.646	2.524	2.057	2.077	2.333 E7	11.27
19) B	Endosulfan Sulfate	2.634	2.980	2.821	2.435	2.315	2.637 E7	10.31
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone	2.471	2.644	2.632	2.323	2.244	2.463 E7	7.29
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD040219CLP.M
 Title : GC Extractables
 Last Update : Mon Apr 01 19:05:44 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD052439.D 1 =PD052435.D 2 =PD052437.D
 4 =PD052441.D 5 =PD052443.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.948	2.135	2.087	1.748	1.782	1.940 E7	9.00

(#) = Out of Range