

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040219\
 Data File : PD052458.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Apr 2019 22:26
 Operator : AJ\SJ
 Sample : K2122-09
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleID :
 BF5D0

Manual Integrations
 APPROVED

Sohil
 4/3/2019 9:08:41 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 05:43:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD040219CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Apr 01 19:09:32 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.220	4.041	17330310	406.6E6	12.218	12.724
27)	SA Decachlor...	7.811	9.153	63482198	413.2E6	42.256m	21.254m#
Target Compounds							
10)	B trans-Chl...	5.060	6.162	1959.1E6	24582.9E6	952.099m	650.193m#
11)	B cis-Chlor...	5.116	6.236	2382.3E6	26755.1E6	1186.530	740.618 #
15)	B Endosulfa...	5.921	6.980	97445103	310.7E6	52.031	10.910 #
19)	B Endosulfa...	6.294	7.324	7314188	103.7E6	4.312m	3.934m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040219\
Data File : PD052458.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Apr 2019 22:26
Operator : AJ\SJ
Sample : K2122-09
Misc :
ALS Vial : 32 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 05:43:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD040219CLP.M
Quant Title : GC Extractables
QLast Update : Mon Apr 01 19:09:32 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



4/3/2019 9:08:41 AM

Manual Integrations
APPROVED

Instrument :
ECD_D
ClientSampleId :
BF5D0