

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040422\
 Data File : PD068906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2022 10:09
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 01:11:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 03:30:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.446	4.173	57569709	455.1E6	47.415	47.603
2)	SA Decachlor...	8.203	9.367	57538669	274.3E6	45.544	48.412
Target Compounds							
2)	A alpha-BHC	3.885	4.569	87317158	663.7E6	48.226	48.222
3)	MA gamma-BHC...	4.169	4.859	81138632	594.7E6	48.593	47.564
4)	MA Heptachlor	4.467	5.387	77134974	560.7E6	46.820	47.179
5)	MB Aldrin	4.714	5.699	76443281	547.3E6	47.833	47.681
6)	B beta-BHC	4.419	5.022	34779825	258.1E6	46.712	47.014
7)	B delta-BHC	4.623	5.248	79684944	548.7E6	48.760	45.631
8)	B Heptachlo...	5.158	6.087	70554297	470.3E6	47.178	45.032
9)	A Endosulfan I	5.497	6.450	67076707	444.1E6	47.938	47.357
10)	B gamma-Chl...	5.384	6.323	72028930	486.9E6	47.483	47.235
11)	B alpha-Chl...	5.443	6.398	71073010	472.4E6	47.552	47.444
12)	B 4,4'-DDE	5.609	6.548	66401586	443.4E6	48.537	47.846
13)	MA Dieldrin	5.741	6.708	69718367	459.1E6	47.748	47.268
14)	MA Endrin	5.999	6.928	63060911	376.9E6	47.871	48.715
15)	B Endosulfa...	6.273	7.133	58616233	355.6E6	47.030	47.041
16)	A 4,4'-DDD	6.126	7.039	57066267	354.7E6	48.393	48.029
17)	MA 4,4'-DDT	6.367	7.342	56486202	357.4E6	46.995	48.648
18)	B Endrin al...	6.442	7.257	45175314	280.1E6	45.505	46.291
19)	B Endosulfa...	6.655	7.482	59484253	352.4E6	46.515	47.919
20)	A Methoxychlor	6.917	7.796	31539171	175.9E6	45.509	48.249
21)	B Endrin ke...	7.149	7.957	69385368	370.5E6	47.132	47.816
22)	Mirex	7.332	8.423	52874570	265.1E6	44.613	48.357

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040422\
Data File : PD068906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2022 10:09
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 01:11:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 03:30:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

