

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040422\
 Data File : PD068915.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2022 14:08
 Operator : AR\AJ
 Sample : N2226-01MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PFS-TPMS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 01:23:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 03:30:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.446	4.172	21541630	187.3E6	17.742	19.587
28)	SA Decachlor...	8.204	9.368	22290851	111.8E6	17.644	19.737
Target Compounds							
2)	A alpha-BHC	3.885	4.569	84870368	665.6E6	46.875	48.361
3)	MA gamma-BHC...	4.169	4.859	78263665	599.6E6	46.871	47.957
4)	MA Heptachlor	4.467	5.386	73896713	575.9E6	44.854	48.460
5)	MB Aldrin	4.714	5.699	76329160	575.9E6	47.762	50.173
6)	B beta-BHC	4.419	5.022	33192929	249.4E6	44.581	45.425
7)	B delta-BHC	4.623	5.248	78430969	591.8E6	47.993	49.210
8)	B Heptachlo...	5.158	6.088	68875619	498.7E6	46.056	47.754
9)	A Endosulfan I	5.497	6.450	63892674	456.3E6	45.662	48.656
10)	B gamma-Chl...	5.384	6.324	70655629	501.5E6	46.578	48.653
11)	B alpha-Chl...	5.444	6.398	69448367	487.8E6	46.465	48.995
12)	B 4,4'-DDE	5.609	6.548	65005052	452.6E6	47.516	48.842
13)	MA Dieldrin	5.741	6.709	71149511	485.4E6	48.728	49.969
14)	MA Endrin	5.999	6.929	64244056	401.3E6	48.769	51.866
15)	B Endosulfa...	6.273	7.134	62321641	394.2E6	50.003	52.148
16)	A 4,4'-DDD	6.126	7.040	54484272	368.1E6	46.203	49.845
17)	MA 4,4'-DDT	6.367	7.343	55032317	358.4E6	45.785	48.784
18)	B Endrin al...	6.443	7.258	47852299	307.4E6	48.202	50.789
19)	B Endosulfa...	6.656	7.482	61828425	364.8E6	48.348	49.602
20)	A Methoxychlor	6.918	7.796	30935503	184.1E6	44.638	50.504
21)	B Endrin ke...	7.149	7.958	67521244	363.4E6	45.865	46.895
22)	Mirex	7.332	8.424	52071576	269.0E6	43.936	49.078

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040422\
Data File : PD068915.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2022 14:08
Operator : AR\AJ
Sample : N2226-01MS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PFS-TPMS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 01:23:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 03:30:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

