

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD040519\  
 Data File : PD052631.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 04 Apr 2019 3:48  
 Operator : AJ\SJ  
 Sample : PB118458BS  
 Misc :  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 PB118523BL

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Apr 04 07:01:58 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD032919.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Mar 29 13:06:45 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2   | ng/ml  | ng/ml   |
|-----------------------------|-----------------|-------|-------|----------|----------|--------|---------|
| -----                       |                 |       |       |          |          |        |         |
| System Monitoring Compounds |                 |       |       |          |          |        |         |
| 1)                          | SA Tetrachlo... | 3.220 | 4.042 | 26319412 | 562.2E6  | 18.861 | 19.397  |
| 28)                         | SA Decachlor... | 7.815 | 9.157 | 27814768 | 360.9E6  | 20.870 | 23.205  |
| Target Compounds            |                 |       |       |          |          |        |         |
| 3)                          | MA gamma-BHC... | 3.916 | 0.000 | 3948645  | 0        | 1.806  | N.D. #  |
| 4)                          | MA Heptachlor   | 0.000 | 5.245 | 0        | 1827111  | N.D.   | 0.046 # |
| 10)                         | B gamma-Chl...  | 0.000 | 6.177 | 0        | 23624336 | N.D.   | 0.738 # |
| 15)                         | B Endosulfa...  | 0.000 | 6.980 | 0        | 2216640  | N.D.   | 0.089 # |
| 25)                         | Chlordane-3     | 0.000 | 6.177 | 0        | 23624336 | N.D.   | 5.354 # |
| -----                       |                 |       |       |          |          |        |         |



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD040519\  
Data File : PD052631.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 04 Apr 2019 3:48  
Operator : AJ\SJ  
Sample : PB118458BS  
Misc :  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
ECD\_D  
ClientSampleId :  
PB118523BL

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Apr 04 07:01:58 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD032919.M  
Quant Title : GC Extractables  
QLast Update : Fri Mar 29 13:06:45 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.220 min  
Delta R.T.: -0.001 min  
Response: 26319412  
Conc: 18.86 ng/ml

Instrument :  
ECD\_D  
ClientSampleId :  
PB118523BL




#1 Tetrachloro-m-xylene

R.T.: 4.042 min  
Delta R.T.: 0.000 min  
Response: 562201532  
Conc: 19.40 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 3.916 min  
Delta R.T.: 0.007 min  
Response: 3948645  
Conc: 1.81 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 0.000 min  
Exp R.T. : 4.723 min  
Response: 0  
Conc: N.D.



#4 Heptachlor


R.T.: 0.000 min  
Exp R.T. : 4.187 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_D  
ClientSampleId :  
PB118523BL




#4 Heptachlor

R.T.: 5.245 min  
Delta R.T.: 0.010 min  
Response: 1827111  
Conc: 0.05 ng/ml




#10 gamma-Chlordane

R.T.: 0.000 min  
Exp R.T. : 5.063 min  
Response: 0  
Conc: N.D.



#10 gamma-Chlordane


R.T.: 6.177 min  
Delta R.T.: 0.012 min  
Response: 23624336  
Conc: 0.74 ng/ml



#15 Endosulfan II


R.T.: 0.000 min  
Exp R.T. : 5.919 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_D  
ClientSampleId :  
PB118523BL




#15 Endosulfan II

R.T.: 6.980 min  
Delta R.T.: 0.004 min  
Response: 2216640  
Conc: 0.09 ng/ml




#25 Chlordane-3

R.T.: 0.000 min  
Exp R.T. : 5.064 min  
Response: 0  
Conc: N.D.



#25 Chlordane-3


R.T.: 6.177 min  
Delta R.T.: 0.011 min  
Response: 23624336  
Conc: 5.35 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.815 min  
Delta R.T.: -0.003 min  
Response: 27814768  
Conc: 20.87 ng/ml

Instrument :  
ECD\_D  
ClientSampleId :  
PB118523BL



#28 Decachlorobiphenyl

R.T.: 9.157 min  
Delta R.T.: 0.000 min  
Response: 360875865  
Conc: 23.21 ng/ml