

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040519\
 Data File : PD052611.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Apr 2019 15:04
 Operator : AJ\SJ
 Sample : PTOXCCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXCCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 06:08:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX032919.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 29 14:07:32 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.221	4.044	68952671	1374.7E6	50.751	49.276
7) SA Decachlor...	7.818	9.161	65872622	816.8E6	49.046	53.321
Target Compounds						
2) Toxaphene-1	5.371	6.448	4234124	70770504	513.490	510.142
3) Toxaphene-2	5.951	6.739	14204034	82110364	518.297	501.909
4) Toxaphene-3	6.190	7.226	13775990	305.7E6	530.017	529.366
5) Toxaphene-4	6.536	7.312	10227964	185.9E6	535.549	544.272
6) Toxaphene-5	6.663	7.729	17151297	128.0E6	522.589	522.298

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040519\
Data File : PD052611.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Apr 2019 15:04
Operator : AJ\SJ
Sample : PTOXCCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 06:08:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX032919.M
Quant Title : GC Extractables
QLast Update : Fri Mar 29 14:07:32 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

