

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040519\
 Data File : PD052599.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Apr 2019 12:16
 Operator : AJ\SJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORCCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 06:50:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 29 13:06:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.221	4.043	73569435	1414.8E6	52.722	48.813
28)	SA Decachlor...	7.818	9.162	67940454	829.5E6	50.977	53.338
Target Compounds							
23)	Chlordane-1	4.042	5.046	32455908	744.0E6	516.864	484.944
24)	Chlordane-2	4.522	5.517	37887507	661.7E6	509.130	472.492
25)	Chlordane-3	5.063	6.167	115.4E6	2039.3E6	521.907	462.124
26)	Chlordane-4	5.118	6.241	110.8E6	2399.2E6	522.259	471.797
27)	Chlordane-5	5.923	7.051	34230405	503.6E6	526.354	545.830



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040519\
Data File : PD052599.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Apr 2019 12:16
Operator : AJ\SJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 06:50:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
Quant Title : GC Extractables
QLast Update : Fri Mar 29 13:06:45 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.221 min
Delta R.T.: 0.000 min
Response: 73569435
Conc: 52.72 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500




#1 Tetrachloro-m-xylene

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 1414785910
Conc: 48.81 ng/ml




#23 Chlordane-1

R.T.: 4.042 min
Delta R.T.: 0.000 min
Response: 32455908
Conc: 516.86 ng/ml



#23 Chlordane-1


R.T.: 5.046 min
Delta R.T.: 0.000 min
Response: 743952836
Conc: 484.94 ng/ml



#24 Chlordane-2


R.T.: 4.522 min
Delta R.T.: 0.000 min
Response: 37887507
Conc: 509.13 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500




#24 Chlordane-2

R.T.: 5.517 min
Delta R.T.: 0.000 min
Response: 661710513
Conc: 472.49 ng/ml




#25 Chlordane-3

R.T.: 5.063 min
Delta R.T.: -0.001 min
Response: 115413169
Conc: 521.91 ng/ml



#25 Chlordane-3


R.T.: 6.167 min
Delta R.T.: 0.000 min
Response: 2039259184
Conc: 462.12 ng/ml



#26 Chlordane-4


R.T.: 5.118 min
Delta R.T.: 0.000 min
Response: 110767304
Conc: 522.26 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500




#26 Chlordane-4

R.T.: 6.241 min
Delta R.T.: 0.000 min
Response: 2399156329
Conc: 471.80 ng/ml




#27 Chlordane-5

R.T.: 5.923 min
Delta R.T.: 0.000 min
Response: 34230405
Conc: 526.35 ng/ml



#27 Chlordane-5


R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 503587834
Conc: 545.83 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.818 min
Delta R.T.: 0.000 min
Response: 67940454
Conc: 50.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500



#28 Decachlorobiphenyl

R.T.: 9.162 min
Delta R.T.: 0.005 min
Response: 829481076
Conc: 53.34 ng/ml