

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040519\
 Data File : PD052640.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Apr 2019 5:56
 Operator : AJ\SJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORCCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 07:05:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 29 13:06:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.220	4.042	72260322	1362.8E6	51.784	47.019
28)	SA Decachlor...	7.815	9.158	65907156	836.3E6	49.452	53.778
Target Compounds							
23)	Chlordane-1	4.041	5.044	32497982	720.0E6	517.534	469.350
24)	Chlordane-2	4.520	5.515	38379865	657.2E6	515.746	469.273
25)	Chlordane-3	5.061	6.164	117.5E6	2104.2E6	531.521	476.848
26)	Chlordane-4	5.116	6.239	111.8E6	2395.7E6	526.993	471.120
27)	Chlordane-5	5.920	7.049	34118258	490.5E6	524.630	531.674



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040519\
Data File : PD052640.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Apr 2019 5:56
Operator : AJ\SJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 07:05:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032919.M
Quant Title : GC Extractables
QLast Update : Fri Mar 29 13:06:45 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.220 min
Delta R.T.: 0.000 min
Response: 72260322
Conc: 51.78 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500




#1 Tetrachloro-m-xylene

R.T.: 4.042 min
Delta R.T.: 0.000 min
Response: 1362793455
Conc: 47.02 ng/ml




#23 Chlordane-1

R.T.: 4.041 min
Delta R.T.: -0.002 min
Response: 32497982
Conc: 517.53 ng/ml



#23 Chlordane-1


R.T.: 5.044 min
Delta R.T.: -0.002 min
Response: 720030090
Conc: 469.35 ng/ml



#24 Chlordane-2


R.T.: 4.520 min
Delta R.T.: -0.003 min
Response: 38379865
Conc: 515.75 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500




#24 Chlordane-2

R.T.: 5.515 min
Delta R.T.: -0.001 min
Response: 657203308
Conc: 469.27 ng/ml




#25 Chlordane-3

R.T.: 5.061 min
Delta R.T.: -0.003 min
Response: 117539184
Conc: 531.52 ng/ml



#25 Chlordane-3


R.T.: 6.164 min
Delta R.T.: -0.002 min
Response: 2104234118
Conc: 476.85 ng/ml



#26 Chlordane-4


R.T.: 5.116 min
Delta R.T.: -0.002 min
Response: 111771444
Conc: 526.99 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500




#26 Chlordane-4

R.T.: 6.239 min
Delta R.T.: -0.002 min
Response: 2395714654
Conc: 471.12 ng/ml




#27 Chlordane-5

R.T.: 5.920 min
Delta R.T.: -0.004 min
Response: 34118258
Conc: 524.63 ng/ml



#27 Chlordane-5


R.T.: 7.049 min
Delta R.T.: -0.001 min
Response: 490527341
Conc: 531.67 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.815 min
Delta R.T.: -0.004 min
Response: 65907156
Conc: 49.45 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500



#28 Decachlorobiphenyl

R.T.: 9.158 min
Delta R.T.: 0.001 min
Response: 836323592
Conc: 53.78 ng/ml