

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040523\
 Data File : PD074593.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2023 09:56
 Operator : AR\AJ
 Sample : PB151880BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS880

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 22:22:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 14 03:05:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.771	48069589	24127522	19.349	19.178
27)	SA Decachlor...	9.094	7.920	108.5E6	45514293	39.632	38.011
Target Compounds							
2)	A alpha-BHC	4.007	3.274	16490020	7460219	4.226	4.175
3)	MA gamma-BHC...	4.341	3.604	15660927	7202601	4.243	4.217
4)	MA Heptachlor	4.932	3.946	17775827	7355637	4.695	4.436
5)	MB Aldrin	5.276	4.226	17960316	7654668	4.914	4.471
6)	B beta-BHC	4.535	3.900	8265411	3817630	4.889	4.712
7)	B delta-BHC	4.783	4.131	18733039	6215917	5.148	3.514 #
8)	B Heptachlo...	5.703	4.728	16628433	6945798	5.065	4.384
9)	A Endosulfan I	6.089	5.099	15301035	6450011	5.129	4.600
10)	B trans-Chl...	5.959	4.979	16038213	7020217	4.999	4.538
11)	B cis-Chlor...	6.039	5.044	16576764	7205642	5.059	4.555
12)	B 4,4'-DDE	6.211	5.233	28106243	12328967	9.383	8.488
13)	MA Dieldrin	6.365	5.364	31318509	13810634	9.872	8.908
14)	MA Endrin	6.596	5.640	24156960	12030572	9.167	8.943
15)	B Endosulfa...	6.814	5.933	27796019	12427099	10.070	9.243
16)	A 4,4'-DDD	6.728	5.787	23874632	10889384	10.102	9.249
17)	MA 4,4'-DDT	7.044	6.039	23781619	10355341	9.477	8.266
18)	B Endrin al...	6.944	6.113	20816504	9466015	9.242	8.611
19)	B Endosulfa...	7.180	6.336	26065917	11599157	10.008	9.215
20)	A Methoxychlor	7.519	6.615	67684146	28783636	49.301	46.796
21)	B Endrin ke...	7.666	6.842	30834199	14610074	10.400	9.602

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040523\
Data File : PD074593.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Apr 2023 09:56
Operator : AR\AJ
Sample : PB151880BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 22:22:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

