

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040622\
 Data File : PD068971.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2022 12:08
 Operator : AR\AJ
 Sample : INDB345
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB3164

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 04:58:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 07:24:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.451	4.174	23613615	222.2E6	17.340	20.303
27)	SA Decachlor...	8.210	9.372	50358481	257.6E6	36.215	40.846
Target Compounds							
5)	MB Aldrin	4.721	5.701	30985428	279.7E6	17.618	20.729
6)	B beta-BHC	4.425	5.024	14948047	127.3E6	17.938	19.930
7)	B delta-BHC	4.629	5.250	31129877	282.6E6	17.606	19.961
8)	B Heptachlo...	5.165	6.089	28827357	266.9E6	17.751	22.157
10)	B trans-Chl...	5.391	6.326	29426029	246.8E6	17.632	20.752
11)	B cis-Chlor...	5.450	6.401	29046903	241.5E6	17.695	21.056
12)	B 4,4'-DDE	5.616	6.551	53322704	426.9E6	35.847	41.782
15)	B Endosulfa...	6.279	7.137	51692472	369.2E6	35.638	41.261
18)	B Endrin al...	6.449	7.260	39566279	280.6E6	36.080	42.251
19)	B Endosulfa...	6.661	7.485	50151774	341.7E6	35.884	41.720
21)	B Endrin ke...	7.155	7.961	57408809	354.0E6	35.980	41.457

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040622\
Data File : PD068971.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2022 12:08
Operator : AR\AJ
Sample : INDB345
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 04:58:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 07:24:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

