

Method Path : P:\HPCHEM1\ECD_D\Method\
 Method File : DTX040718.M
 Title : GC Extractables
 Last Update : Sat Apr 07 11:40:20 2018
 Response Via : Initial Calibration

Calibration Files

500 =PD046825.D 1000 =PD046823.D 750 =PD046824.D
 250 =PD046826.D =

Compound		500	1000	750	250	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.205	1.264	1.205	1.212	1.222	E6	2.33
2)	Toxaphene-1	9.314	9.892	9.066	9.480	9.438	E3	3.68
3)	Toxaphene-2	1.021	1.013	0.970	1.044	1.012	E4	3.03
4)	Toxaphene-3	3.567	3.716	3.560	4.514	3.839	E3	11.86
5)	Toxaphene-4	8.076	8.863	8.391	8.357	8.422	E3	3.88
6)	Toxaphene-5	3.500	3.585	3.524	3.181	3.447	E3	5.26
7) SA	Decachlorobiphenyl	1.290	1.305	1.260	1.306	1.290	E6	1.67

Signal #2 Calibration Files

500 =PD046825.D 1000 =PD046823.D 750 =PD046824.D
 250 =PD046826.D =

Compound		500	1000	750	250	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.997	1.953	1.922	2.147	2.005	E7	4.99
2)	Toxaphene-1	5.816	5.500	5.603	6.013	5.733	E4	3.98
3)	Toxaphene-2	1.890	1.811	1.796	2.026	1.881	E5	5.58
4)	Toxaphene-3	5.368	5.238	5.176	5.938	5.430	E5	6.41
5)	Toxaphene-4	3.405	3.282	3.267	3.576	3.383	E5	4.22
6)	Toxaphene-5	1.225	1.244	1.178	1.355	1.250	E5	5.99
7) SA	Decachlorobiphenyl	1.682	1.594	1.591	1.815	1.671	E7	6.31

(#) = Out of Range