

Method Path : P:\HPCHEM1\ECD_D\Method\
 Method File : DTX040718.M
 Title : GC Extractables
 Last Update : Sat Apr 07 06:34:55 2018
 Response Via : Initial Calibration

Calibration Files

500 =PD046825.D 1000 =PD046823.D 750 =PD046824.D
 250 =PD046826.D =

	Compound	500	1000	750	250	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.205	1.264	1.205	1.212	1.222	E6	2.33
2)	Toxaphene-1	9.314	9.892	9.066	9.480	9.438	E3	3.68
3)	Toxaphene-2	0.932	1.023	0.964	0.918	0.959	E4	4.85
4)	Toxaphene-3	2.530	2.698	2.587	2.507	2.581	E4	3.30
5)	Toxaphene-4	2.097	2.222	2.180	2.053	2.138	E4	3.60
6)	Toxaphene-5	2.856	2.959	2.840	2.698	2.839	E4	3.78
7) SA	Decachlorobiphenyl	1.290	1.305	1.260	1.306	1.290	E6	1.67

Signal #2 Calibration Files

500 =PD046825.D 1000 =PD046823.D 750 =PD046824.D
 250 =PD046826.D =

	Compound	500	1000	750	250	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.997	1.953	1.922	2.147	2.005	E7	4.99
2)	Toxaphene-1	1.094	1.085	1.018	1.224	1.105	E5	7.81
3)	Toxaphene-2	1.317	1.278	1.225	1.407	1.307	E5	5.88
4)	Toxaphene-3	1.890	1.811	1.796	2.070	1.892	E5	6.64
5)	Toxaphene-4	5.368	5.238	5.176	5.938	5.430	E5	6.41
6)	Toxaphene-5	2.077	2.026	2.233	2.267	2.151	E5	5.45
7) SA	Decachlorobiphenyl	1.682	1.594	1.591	1.815	1.671	E7	6.31

(#) = Out of Range