



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Dewberry Engineers Inc.

Project Name: ORU 129 Indian Hollow Road

Project # N/A

Chemtech Project # M1831

Test Name: EPH

A. Number of Samples and Date of Receipt:

9 Solid samples were received on 03/25/2021.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: EPH, PCB and SVOC-PAH. This data package contains results for EPH.

C. Analytical Techniques:

The analysis were performed on instrument FID_C. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analyses were performed on instrument FID_D. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analysis of EPHs was based on method NJEPH and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for B2B[1-chlorooctadecane - 148%].

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

Sample B2B was diluted due to high concentration.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

Hard copy data reported new MDL and fax was reported old mdl so Hard copy and Fax MDL not match due to data was processes again and some positive hit also not match with fax data.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____