



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

GFE LLC

Project Name: 30-75 21st St., Astoria, NY

Project # N/A

Chemtech Project # M1836

Test Name: Pesticide-TCL

A. Number of Samples and Date of Receipt:

16 Solid samples were received on 03/25/2021.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Cyanide, Mercury, Metals Group1, Metals ICP-Group1, PCB, Pesticide-TCL, SVOC-SIMGroup1, SVOCMS Group1 and VOCMS Group1. This data package contains results for Pesticide-TCL.

C. Analytical Techniques:

The analyses were performed on instrument GCECD_D. The front column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HM-G017-11 . The rear column is ZBMR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 7HM-G016-17. The analysis was performed on instrument ECD_L. The front column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11 The rear column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 7HM-G016-17. .The analysis of Pesticide-TCLs was based on method 8081B and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID PD061875.D met the requirements except for alpha-BHC,delta-BHC,gamma-BHC (Lindane),Heptachlor is failing in 1st column however associated QC having in QC limits.

The Continuous Calibration File ID PD061890.D met the requirements except for 4,4-DDD,4,4-DDE,Aldrin,alpha-BHC,alpha-Chlordane,beta-BHC,delta-BHC, Dieldrin, Endosulfan I, Endrin aldehyde,gamma-BHC (Lindane),gamma-Chlordane,Heptachlor epoxide,Heptachlor,Methoxychlor,Tetrachloro-m-xylene is failing in 1st column however associated QC having in QC limits.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____