

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040724\
 Data File : PD082299.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2024 15:21
 Operator : AR\AJ
 Sample : P2018-08MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BH5F0MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/10/2024
 Supervised By : Ankita Jodhani 04/10/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 08 01:27:37 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032924CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Mar 30 04:18:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.564	2.761	10517100	1104.4E6	7.112m	330.126m#
27)	SA Decachlor...	9.120	7.909	11684781	10860705	6.288	4.117m#
Target Compounds							
2)	A alpha-BHC	4.024	3.278	63020072	153.6E6	25.267	30.340
3)	MA gamma-BHC...	4.356	3.606	57150837	169.1E6	23.877m	36.616 #
4)	MA Heptachlor	4.951	3.944	131.5E6	140.7E6	53.071	32.825m#
5)	MB Aldrin	5.293	4.223	64049079	128.1E6	27.357	30.557m
6)	B beta-BHC	4.552	3.902	25686799	57663059	24.520m	29.782
7)	B delta-BHC	4.800	4.131	35277764	73706806	15.145m	16.423m
8)	B Heptachlo...	5.720	4.727	54232052	112.2E6	24.701	30.248
9)	A Endosulfan I	6.105	5.098	69265695	122.4E6	33.145m	36.041
10)	B trans-Chl...	5.976	4.976	64752314	111.9E6	31.115m	31.099
11)	B cis-Chlor...	6.057	5.041	52327315	123.0E6	24.692	33.313 #
12)	B 4,4'-DDE	6.227	5.228	117.4E6	212.9E6	57.626m	62.471m
13)	MA Dieldrin	6.384	5.359	131.9E6	246.1E6	61.232	67.044m
14)	MA Endrin	6.615	5.635	118.4E6	209.1E6	66.706	67.547
15)	B Endosulfa...	6.831	5.927	72446488	178.8E6	40.953m	59.012m#
16)	A 4,4'-DDD	6.746	5.782	90762367	121.4E6	60.546m	45.058m#
17)	MA 4,4'-DDT	7.063	6.032	94225338	165.7E6	60.581	59.988m
18)	B Endrin al...	6.964	6.108	77986288	130.3E6	58.838	57.740
19)	B Endosulfa...	7.198	6.330	78882829	153.4E6	48.389	53.815
20)	A Methoxychlor	7.538	6.607	216.9E6	414.0E6	264.605	296.992m
21)	B Endrin ke...	7.685	6.837	87898876	193.5E6	51.405m	60.500

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040724\
Data File : PD082299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2024 15:21
Operator : AR\AJ
Sample : P2018-08MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:27:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032924CLP.M
Quant Title : GC Extractables
QLast Update : Sat Mar 30 04:18:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
Client Sampled :
BHSFOMS

Manual Integrations
APPROVED

Reviewed By: Abdul Mirza 04/10/2024
Supervised By: Ankita Jodhani 04/10/2024