

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040821\
 Data File : PD062063.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2021 13:18
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 09 02:36:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032921.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 30 04:47:56 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.484	3.980	39791767	405.9E6	47.524	46.713
28)	SA Decachlor...	8.255	9.038	56425391	386.2E6	45.858	40.906
Target Compounds							
2)	A alpha-BHC	3.924	4.369	61297693	652.6E6	48.533	48.521
3)	MA gamma-BHC...	4.209	4.653	57956933	591.7E6	48.815	47.435
4)	MA Heptachlor	4.512	5.161	59912953	571.3E6	51.464	43.843
5)	MB Aldrin	4.761	5.466	58808135	574.1E6	49.001	46.134
6)	B beta-BHC	4.458	4.822	27048349	242.2E6	49.601	46.885
7)	B delta-BHC	4.665	5.039	59586456	578.0E6	49.050	43.698
8)	B Heptachlo...	5.205	5.851	55227744	515.2E6	48.177	43.510
9)	A Endosulfan I	5.546	6.208	52304183	498.1E6	47.631	45.952
10)	B gamma-Chl...	5.433	6.085	56263193	544.4E6	48.033	45.924
11)	B alpha-Chl...	5.492	6.158	55643193	519.0E6	47.885	45.525
12)	B 4,4'-DDE	5.654	6.315	53692300	524.8E6	47.783	45.541
13)	MA Dieldrin	5.790	6.465	55257928	521.7E6	47.805	45.197
14)	MA Endrin	6.049	6.682	47403671	415.0E6	46.135	41.940
15)	B Endosulfa...	6.321	6.892	47481939	417.6E6	46.329	43.986
16)	A 4,4'-DDD	6.172	6.805	46480274	424.7E6	49.945	44.851
17)	MA 4,4'-DDT	6.414	7.103	44128922	421.2E6	48.993	44.102
18)	B Endrin al...	6.490	7.015	39228848	346.6E6	45.916	44.070
19)	B Endosulfa...	6.704	7.238	49167732	426.5E6	45.566	42.551
20)	A Methoxychlor	6.960	7.561	25685290	196.9E6	54.360	43.673
21)	B Endrin ke...	7.198	7.708	58501689	429.7E6	45.839	42.362
22)	Mirex	7.388	8.163	43272914	266.4E6	51.836	41.336

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040821\
Data File : PD062063.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2021 13:18
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 09 02:36:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032921.M
Quant Title : GC Extractables
QLast Update : Tue Mar 30 04:47:56 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

