

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040919\
 Data File : PD052779.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Apr 2019 8:58
 Operator : AJ\SJ
 Sample : PB118631BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 01:36:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD040819CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Apr 09 07:32:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.221	4.043	32441344	678.8E6	21.307	21.451
27)	SA Decachlor...	7.817	9.163	55716761	600.7E6	39.516	41.661
Target Compounds							
6)	B beta-BHC	0.000	4.933f	0 25974729	N.D.	1.328	
15)	B Endosulfa...	0.000	6.989	0 16429609	N.D.	0.624	


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD040919\
Data File : PD052779.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Apr 2019 8:58
Operator : AJ\SJ
Sample : PB118631BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 01:36:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD040819CLP.M
Quant Title : GC Extractables
QLast Update : Tue Apr 09 07:32:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.221 min
Delta R.T.: 0.000 min
Response: 32441344
Conc: 21.31 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 678840317
Conc: 21.45 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.155 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 4.933 min
Delta R.T.: 0.040 min
Response: 25974729
Conc: 1.33 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 5.919 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 6.989 min
Delta R.T.: 0.010 min
Response: 16429609
Conc: 0.62 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.817 min
Delta R.T.: 0.000 min
Response: 55716761
Conc: 39.52 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.163 min
Delta R.T.: 0.001 min
Response: 600695996
Conc: 41.66 ng/ml