

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041023\
 Data File : PD074631.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2023 10:29
 Operator : AR\AJ
 Sample : PSTDCCC50
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC50

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 22:38:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031423.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 15 12:23:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.558	2.770	102.8E6	56837340	44.088	46.703
28) SA Decachlor...	9.104	7.925	127.3E6	58202212	47.615	47.438
Target Compounds						
2) A alpha-BHC	4.015	3.275	180.4E6	87710893	46.493	47.197
3) MA gamma-BHC...	4.349	3.605	161.3E6	81226969	45.974	47.413
4) MA Heptachlor	4.941	3.947	171.7E6	78717152	50.111	46.144
5) MB Aldrin	5.284	4.228	164.7E6	80943452	48.915	47.978
6) B beta-BHC	4.544	3.901	66761799	34943057	45.174	46.276
7) B delta-BHC	4.793	4.132	161.7E6	81159168	58.927	47.681
8) B Heptachlo...	5.711	4.731	146.3E6	72955661	52.021	47.525
9) A Endosulfan I	6.098	5.101	139.9E6	69251922	49.219	48.121
10) B gamma-Chl...	5.968	4.982	150.8E6	74569986	50.261	48.027
11) B alpha-Chl...	6.048	5.046	149.7E6	74157295	49.635	48.154
12) B 4,4'-DDE	6.219	5.235	139.0E6	68719533	50.430	48.698
13) MA Dieldrin	6.374	5.367	146.8E6	73072512	49.755	48.597
14) MA Endrin	6.604	5.642	122.4E6	63972349	49.262	48.947
15) B Endosulfa...	6.823	5.936	119.3E6	60368234	46.325	47.977
16) A 4,4'-DDD	6.737	5.790	113.8E6	57229506	50.753	49.027
17) MA 4,4'-DDT	7.052	6.042	126.4E6	58138986	49.588	47.344
18) B Endrin al...	6.953	6.116	96490256	48249615	48.270	47.512
19) B Endosulfa...	7.188	6.339	122.1E6	60615395	49.464	49.650
20) A Methoxychlor	7.528	6.618	68397534	30220844	48.884	44.664
21) B Endrin ke...	7.675	6.845	139.9E6	72195216	50.264	48.943
22) Mirex	8.150	7.029	104.3E6	56792665	49.126	49.999

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041023\
Data File : PD074631.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2023 10:29
Operator : AR\AJ
Sample : PSTDCCC50
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC50

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 22:38:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031423.M
Quant Title : GC Extractables
QLast Update : Wed Mar 15 12:23:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

