

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041119\
 Data File : PD052865.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Apr 2019 18:02
 Operator : AJ\SJ
 Sample : K2218-10
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 01:04:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD040819CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Apr 09 07:32:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.218	4.041	26756647	515.5E6	17.574	16.290
27)	SA Decachlor...	7.812	9.158	43044007	465.5E6	30.528	32.285
Target Compounds							
2)	A alpha-BHC	0.000	4.405f	0 44574895	N.D.	0.937	
6)	B beta-BHC	0.000	4.933f	0 21068396	N.D.	1.077	
8)	B Heptachlo...	0.000	5.908	0 97845507	N.D.	2.762	
9)	A Endosulfan I	0.000	6.310	0 5673047	N.D.	0.173	
15)	B Endosulfa...	0.000	6.978	0 36393370	N.D.	1.382	


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041119\
Data File : PD052865.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Apr 2019 18:02
Operator : AJ\SJ
Sample : K2218-10
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 01:04:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD040819CLP.M
Quant Title : GC Extractables
QLast Update : Tue Apr 09 07:32:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.218 min
Delta R.T.: -0.003 min
Response: 26756647
Conc: 17.57 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.041 min
Delta R.T.: -0.002 min
Response: 515511323
Conc: 16.29 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 3.640 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 4.405 min
Delta R.T.: -0.031 min
Response: 44574895
Conc: 0.94 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T.: 4.155 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#6 beta-BHC

R.T.: 4.933 min
Delta R.T.: 0.039 min
Response: 21068396
Conc: 1.08 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T.: 4.845 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.908 min
Delta R.T.: -0.023 min
Response: 97845507
Conc: 2.76 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 5.165 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#9 Endosulfan I

R.T.: 6.310 min
Delta R.T.: 0.019 min
Response: 5673047
Conc: 0.17 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 5.919 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.978 min
Delta R.T.: -0.001 min
Response: 36393370
Conc: 1.38 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.812 min
Delta R.T.: -0.006 min
Response: 43044007
Conc: 30.53 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 9.158 min
Delta R.T.: -0.004 min
Response: 465507359
Conc: 32.28 ng/ml