

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD041124CLP-TCLP.M
 Title : GC Extractables
 Last Update : Wed Apr 10 04:31:43 2024
 Response Via : Initial Calibration

Calibration Files

3	=PD082354.D	1	=PD082352.D	2	=PD082353.D
4	=PD082355.D	5	=PD082356.D	6	=PD082357.D

Compound		3	1	2	4	5	6	Avg	%RSD	

1) SA	Tetrachloro-m...	1.481	1.593	1.471	1.545	1.459	1.441	1.498	E6	3.90
2) A	alpha-BHC							0.000		-1.00
3) MA	gamma-BHC (Li...							0.000		-1.00
4) MA	Heptachlor							0.000		-1.00
5) MB	Aldrin							0.000		-1.00
6) B	beta-BHC							0.000		-1.00
7) B	delta-BHC							0.000		-1.00
8) B	Heptachlor ep...							0.000		-1.00
9) A	Endosulfan I							0.000		-1.00
10) B	trans-Chlordane							0.000		-1.00
11) B	cis-Chlordane							0.000		-1.00
12) B	4,4'-DDE							0.000		-1.00
13) MA	Dieldrin							0.000		-1.00
14) MA	Endrin							0.000		-1.00
15) B	Endosulfan II							0.000		-1.00
16) A	4,4'-DDD							0.000		-1.00
17) MA	4,4'-DDT							0.000		-1.00
18) B	Endrin aldehyde							0.000		-1.00
19) B	Endosulfan Su...							0.000		-1.00
20) A	Methoxychlor							0.000		-1.00
21) B	Endrin ketone							0.000		-1.00
22)	Toxaphene-1	1.566	1.601	1.565	1.589	1.570	1.560	1.575	E4	1.02
23)	Toxaphene-2	1.802	1.855	1.739	1.910	1.907	1.951	1.861	E4	4.25
24)	Toxaphene-3	5.068	4.919	4.972	5.396	5.502	5.541	5.233	E4	5.32
25)	Toxaphene-4	3.765	3.696	3.666	4.022	4.076	4.166	3.899	E4	5.52
26)	Toxaphene-5	2.684	2.677	2.626	2.940	2.824	2.961	2.785	E4	5.17
27) SA	Decachlorobip...	1.212	1.282	1.248	1.209	1.140	1.089	1.197	E6	5.93

Signal #2 Calibration Files

3	=PD082354.D	1	=PD082352.D	2	=PD082353.D
4	=PD082355.D	5	=PD082356.D	6	=PD082357.D

Compound		3	1	2	4	5	6	Avg	%RSD	
1) SA	Tetrachloro-m...	3.409	3.400	3.448	3.448	3.380	3.276	3.394	E6	1.88
2) A	alpha-BHC							0.000		-1.00
3) MA	gamma-BHC (Li...							0.000		-1.00
4) MA	Heptachlor							0.000		-1.00
5) MB	Aldrin							0.000		-1.00
6) B	beta-BHC							0.000		-1.00
7) B	delta-BHC							0.000		-1.00
8) B	Heptachlor ep...							0.000		-1.00
9) A	Endosulfan I							0.000		-1.00
10) B	trans-Chlordane							0.000		-1.00
11) B	cis-Chlordane							0.000		-1.00
12) B	4,4'-DDE							0.000		-1.00
13) MA	Dieldrin							0.000		-1.00
14) MA	Endrin							0.000		-1.00
15) B	Endosulfan II							0.000		-1.00
16) A	4,4'-DDD							0.000		-1.00
17) MA	4,4'-DDT							0.000		-1.00
18) B	Endrin aldehyde							0.000		-1.00
19) B	Endosulfan Su...							0.000		-1.00
20) A	Methoxychlor							0.000		-1.00
21) B	Endrin ketone							0.000		-1.00
22)	Toxaphene-1	2.123	2.011	2.079	2.201	2.207	2.203	2.137	E4	3.79

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD041124CLP-TCLP.M
 Title : GC Extractables
 Last Update : Wed Apr 10 04:31:43 2024
 Response Via : Initial Calibration

Calibration Files

3	=PD082354.D	1	=PD082352.D	2	=PD082353.D
4	=PD082355.D	5	=PD082356.D	6	=PD082357.D

	Compound	3	1	2	4	5	6	Avg		%RSD
23)	Toxaphene-2	2.301	2.199	2.318	2.327	2.264	2.205	2.269	E4	2.48
24)	Toxaphene-3	7.815	7.167	7.567	8.605	8.566	8.295	8.003	E4	7.26
25)	Toxaphene-4	1.077	1.007	1.084	1.128	1.112	1.056	1.077	E5	3.98
26)	Toxaphene-5	4.224	3.753	4.254	4.528	4.434	4.286	4.246	E4	6.32
27) SA	Decachlorobip...	1.911	1.859	1.916	1.956	1.887	1.833	1.894	E6	2.31

(#) = Out of Range