

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD041124CLP-TCLP.M
 Title : GC Extractables
 Last Update : Wed Apr 10 06:11:50 2024
 Response Via : Initial Calibration

Calibration Files

3 =PD082362.D 1 =PD082358.D 2 =PD082360.D
 4 =PD082364.D 5 =PD082366.D 6 =PD082368.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	1.613	1.541	1.444	1.571	1.480	1.550	1.533 E6	4.01
2) A alpha-BHC	2.682	2.425	2.196	2.778	2.771	3.057	2.652 E6	11.37
3) MA gamma-BHC (Li...	2.565	2.354	2.179	2.631	2.681	2.874	2.547 E6	9.69
4) MA Heptachlor	2.643	2.562	2.750	2.605	2.688	2.703	2.658 E6	2.58
5) MB Aldrin							0.000	-1.00
6) B beta-BHC							0.000	-1.00
7) B delta-BHC							0.000	-1.00
8) B Heptachlor ep...							0.000	-1.00
9) A Endosulfan I	2.207	2.081	2.076	2.099	2.145	2.107	2.119 E6	2.33
10) B trans-Chlordane							0.000	-1.00
11) B cis-Chlordane							0.000	-1.00
12) B 4,4'-DDE							0.000	-1.00
13) MA Dieldrin	2.316	2.111	2.159	2.284	2.370	2.366	2.268 E6	4.79
14) MA Endrin	1.896	1.721	1.757	1.880	1.935	1.902	1.849 E6	4.73
15) B Endosulfan II							0.000	-1.00
16) A 4,4'-DDD	1.647	1.438	1.488	1.642	1.686	1.708	1.602 E6	6.95
17) MA 4,4'-DDT	1.666	1.481	1.571	1.706	1.748	1.765	1.656 E6	6.67
18) B Endrin aldehyde							0.000	-1.00
19) B Endosulfan Su...							0.000	-1.00
20) A Methoxychlor	8.759	8.666	8.824	8.613	8.365	8.115	8.557 E5	3.13
21) B Endrin ketone							0.000	-1.00
22) Toxaphene-1							0.000	-1.00
23) Toxaphene-2							0.000	-1.00
24) Toxaphene-3							0.000	-1.00
25) Toxaphene-4							0.000	-1.00
26) Toxaphene-5							0.000	-1.00
27) SA Decachlorobip...	1.978	2.095	2.077	1.875	1.832	1.749	1.934 E6	7.17

Signal #2 Calibration Files

3 =PD082362.D 1 =PD082358.D 2 =PD082360.D
 4 =PD082364.D 5 =PD082366.D 6 =PD082368.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	3.483	3.466	3.539	3.465	3.493	3.376	3.470 E6	1.55
2) A alpha-BHC	5.420	4.957	5.259	5.624	5.861	5.774	5.482 E6	6.20
3) MA gamma-BHC (Li...	4.979	4.536	4.877	5.134	5.304	5.204	5.005 E6	5.53
4) MA Heptachlor	4.831	4.582	4.796	4.902	5.019	4.860	4.832 E6	2.99
5) MB Aldrin							0.000	-1.00
6) B beta-BHC							0.000	-1.00
7) B delta-BHC							0.000	-1.00
8) B Heptachlor ep...							0.000	-1.00
9) A Endosulfan I	3.917	3.747	3.597	3.954	3.950	3.907	3.845 E6	3.73
10) B trans-Chlordane							0.000	-1.00
11) B cis-Chlordane							0.000	-1.00
12) B 4,4'-DDE							0.000	-1.00
13) MA Dieldrin	4.389	4.061	4.310	4.456	4.523	4.326	4.344 E6	3.69
14) MA Endrin	3.750	3.528	3.744	3.825	3.865	3.607	3.720 E6	3.47
15) B Endosulfan II							0.000	-1.00
16) A 4,4'-DDD	3.170	2.883	3.113	3.258	3.280	3.324	3.171 E6	5.07
17) MA 4,4'-DDT	3.383	3.002	3.300	3.541	3.634	3.464	3.387 E6	6.55
18) B Endrin aldehyde							0.000	-1.00
19) B Endosulfan Su...							0.000	-1.00
20) A Methoxychlor	1.776	1.780	1.800	1.692	1.549	1.358	1.659 E6	10.50
21) B Endrin ketone							0.000	-1.00
22) Toxaphene-1							0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD041124CLP-TCLP.M
Title : GC Extractables
Last Update : Wed Apr 10 06:11:50 2024
Response Via : Initial Calibration

Calibration Files

3	=PD082362.D	1	=PD082358.D	2	=PD082360.D
4	=PD082364.D	5	=PD082366.D	6	=PD082368.D

	Compound	3	1	2	4	5	6	Avg	%RSD
23)	Toxaphene-2							0.000	-1.00
24)	Toxaphene-3							0.000	-1.00
25)	Toxaphene-4							0.000	-1.00
26)	Toxaphene-5							0.000	-1.00
27) SA	Decachlorobip...	3.321	3.394	3.400	3.256	3.251	3.114	3.289 E6	3.26

(#) = Out of Range