

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041222\
 Data File : PD069071.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2022 12:07
 Operator : AR\AJ
 Sample : PB144002BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS002

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/12/2022
 Supervised By : mohammad ahmed 04/12/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 00:49:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 07:24:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.446	4.172	19999955	195.2E6	14.687	17.835
2)	SA Decachlor...	8.202	9.366	44579434	239.3E6	32.059	37.942
Target Compounds							
2)	A alpha-BHC	3.885	4.568	6422170	66910053	3.355	4.279 #
3)	MA gamma-BHC...	4.168	4.858	6972487	67881978	3.841	4.682
4)	MA Heptachlor	4.466	5.386	6934419	70130923	3.871	5.095 #
5)	MB Aldrin	4.713	5.698	6476267	67771557	3.682	5.022 #
6)	B beta-BHC	4.418	5.021	3558714	32193522	4.271	5.041
7)	B delta-BHC	4.622	5.247	4850949	45161285	2.743	3.190
8)	B Heptachlo...	5.157	6.086	6718413	65017404	4.137	5.398 #
9)	A Endosulfan I	5.495	6.449	6102362	56328657	3.964	5.176 #
10)	B trans-Chl...	5.383	6.323	6728395	62215466	4.032	5.231 #
11)	B cis-Chlor...	5.442	6.398	6653417	60221966	4.053	5.251 #
12)	B 4,4'-DDE	5.608	6.547	11613593	108.6E6	7.807	10.630 #
13)	MA Dieldrin	5.740	6.707	12442368	114.9E6	7.449	10.336 #
14)	MA Endrin	5.998	6.928	11581330	87545892	8.192	10.355 #
15)	B Endosulfa...	6.271	7.131	11005488	97035094	7.588m	10.843m#
16)	A 4,4'-DDD	6.125	7.039	9819418	83665318	7.470	9.972 #
17)	MA 4,4'-DDT	6.366	7.342	9413861	81293554	7.049	10.315 #
18)	B Endrin al...	6.441	7.257	10307245	79004786	9.399	11.898 #
19)	B Endosulfa...	6.654	7.481	10263516	80327321	7.344	9.808 #
20)	A Methoxychlor	6.916	7.794	30109637	201.1E6	42.538	57.410 #
21)	B Endrin ke...	7.147	7.956	13088441	92476230	8.203	10.830 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041222\
Data File : PD069071.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2022 12:07
Operator : AR\AJ
Sample : PB144002BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 00:49:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 07:24:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Miza 04/12/2022
Supervised By: mohammad ahmed 04/12/2022

Manual Integrations
APPROVED

Instrument :
ECD_D
ClientSampled :
PLCS002