

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041222\
 Data File : PD069083.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2022 19:06
 Operator : AR\AJ
 Sample : INDA349
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3171

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 00:54:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 07:24:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.446	4.172	23057307	228.0E6	16.932	20.834
27)	SA Decachlor...	8.203	9.367	47892365	247.1E6	34.441	39.175
Target Compounds							
2)	A alpha-BHC	3.885	4.568	31949227	314.8E6	16.690	20.132
3)	MA gamma-BHC...	4.168	4.859	31740262	296.1E6	17.486	20.422
4)	MA Heptachlor	4.466	5.386	30555724	288.4E6	17.058	20.955
9)	A Endosulfan I	5.497	6.450	26038622	224.8E6	16.913	20.659
13)	MA Dieldrin	5.740	6.708	55478493	457.0E6	33.212	41.131
14)	MA Endrin	5.998	6.928	49723347	359.6E6	35.170	42.531
16)	A 4,4'-DDD	6.126	7.040	43542914	345.1E6	33.127	41.125
17)	MA 4,4'-DDT	6.366	7.342	43783165	335.5E6	32.787	42.572 #
20)	A Methoxychlor	6.917	7.795	122.7E6	768.0E6	173.307	219.222 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041222\
Data File : PD069083.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2022 19:06
Operator : AR\AJ
Sample : INDA349
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 00:54:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 07:24:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

