

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041223\
 Data File : PD074666.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2023 15:09
 Operator : AR\AJ
 Sample : PEM044
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM055

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 22:26:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 14 03:05:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.549	2.770	46357529	25132147	18.660	19.976
27) SA Decachlor...	9.093	7.921	61110819	26012429	22.328	21.724
Target Compounds						
2) A alpha-BHC	4.005	3.273	36630281	17877682	9.387	10.004
3) MA gamma-BHC...	4.339	3.603	36187668	16820787	9.804	9.848
6) B beta-BHC	4.533	3.899	17241481	8551377	10.199	10.555
12) B 4,4'-DDE	6.210	5.236	378222	417503	0.126	0.287 #
14) MA Endrin	6.595	5.640	131.7E6	66123085	49.978	49.152
16) A 4,4'-DDD	6.728	5.788	3479022	1752135	1.472	1.488
17) MA 4,4'-DDT	7.044	6.040	282.8E6	131.2E6	112.689	104.717
18) B Endrin al...	6.943	6.114	4577530	2757126	2.032	2.508
20) A Methoxychlor	7.519	6.616	379.3E6	164.6E6	276.290	267.603
21) B Endrin ke...	7.665	6.842	9328972	4896267	3.146	3.218

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041223\
Data File : PD074666.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2023 15:09
Operator : AR\AJ
Sample : PEM044
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 22:26:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

