

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041224\
 Data File : PD082389.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2024 14:02
 Operator : AR\AJ
 Sample : INDA302
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3214

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 20:55:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041024CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 10 02:08:54 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.566	2.777	32760451	73525797	21.603	21.326
27)	SA Decachlor...	9.121	7.910	82144577	136.2E6	42.006	41.844
Target Compounds							
2)	A alpha-BHC	4.024	3.277	54986604	116.0E6	21.405	21.603
3)	MA gamma-BHC...	4.358	3.606	52880522	107.0E6	21.269	21.330
4)	MA Heptachlor	4.950	3.945	53978338	100.4E6	21.371	21.463
9)	A Endosulfan I	6.109	5.096	44547349	79547852	20.617	20.714
13)	MA Dieldrin	6.385	5.361	91637685	181.8E6	41.593	43.155
14)	MA Endrin	6.616	5.636	77036689	154.9E6	42.019	42.704
16)	A 4,4'-DDD	6.748	5.783	64957507	131.7E6	41.229	42.751
17)	MA 4,4'-DDT	7.063	6.034	67711996	140.6E6	41.398	42.837
20)	A Methoxychlor	7.539	6.609	180.7E6	358.2E6	205.090	213.891

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041224\
Data File : PD082389.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2024 14:02
Operator : AR\AJ
Sample : INDA302
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 20:55:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041024CLP.M
Quant Title : GC Extractables
QLast Update : Wed Apr 10 02:08:54 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

