

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041519\
 Data File : PD052931.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Apr 2019 12:30
 Operator : AJ\SJ
 Sample : INDA302
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA302

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 16 05:30:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD040819CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Apr 09 07:32:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.220	4.043	28010342	626.2E6	18.397	19.787
27)	SA Decachlor...	7.818	9.166	55767564	658.5E6	39.552	45.667
Target Compounds							
2)	A alpha-BHC	3.640	4.436	43057225	946.9E6	17.846	19.909
3)	MA gamma-BHC...	3.908	4.723	41734354	848.2E6	17.093	19.740
4)	MA Heptachlor	4.185	5.236	43487898	859.5E6	17.942	19.696
9)	A Endosulfan I	5.166	6.292	36759403	649.7E6	17.977	19.836
13)	MA Dieldrin	5.401	6.551	82683463	1282.9E6	35.556	39.336
14)	MA Endrin	5.648	6.770	70783217	1004.7E6	35.599	39.749
16)	A 4,4'-DDD	5.789	6.891	65496291	986.9E6	36.412	39.546
17)	MA 4,4'-DDT	6.020	7.192	55392331	745.9E6	33.528	37.929
20)	A Methoxychlor	6.566	7.652	176.0E6	1990.5E6	177.766	209.638

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041519\
Data File : PD052931.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Apr 2019 12:30
Operator : AJ\SJ
Sample : INDA302
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
INDA302

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 16 05:30:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD040819CLP.M
Quant Title : GC Extractables
QLast Update : Tue Apr 09 07:32:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

