

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041720\
 Data File : PD057788.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2020 16:06
 Operator : AJ\MA
 Sample : L2245-30 20X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 TP-28

Manual Integrations
 APPROVED

mohammad
 4/21/2020 8:26:05 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 18 01:03:53 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX032720.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 27 15:53:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.228	3.896	314338	425216	0.441	0.532
7) SA Decachlor...	7.871	8.897	778545	938390	1.007	1.045
Target Compounds						
5) Toxaphene-4	6.373	7.108	278120	1200543	45.197	71.753m#
6) Toxaphene-5	6.652	7.516	68362	268413	13.510	16.452m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041720\
Data File : PD057788.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2020 16:06
Operator : AJ\MA
Sample : L2245-30 20X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 01:03:53 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX032720.M
Quant Title : GC Extractables
QLast Update : Fri Mar 27 15:53:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm




4/21/2020 8:26:05 AM

Manual Integrations
APPROVED

mohammad

Instrument :
ECD_D
ClientSample :
TP-28




#1 Tetrachloro-m-xylene

R.T.: 3.228 min
Delta R.T.: 0.001 min
Response: 314338
Conc: 0.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
TP-28


Manual Integrations
APPROVED

mohammad
4/21/2020 8:26:05 AM




#1 Tetrachloro-m-xylene

R.T.: 3.896 min
Delta R.T.: 0.000 min
Response: 425216
Conc: 0.53 ng/ml




#5 Toxaphene-4

R.T.: 6.373 min
Delta R.T.: 0.003 min
Response: 278120
Conc: 45.20 ng/ml



#5 Toxaphene-4

R.T.: 7.108 min
Delta R.T.: -0.001 min
Response: 1200543
Conc: 71.75 ng/ml m




#6 Toxaphene-5

R.T.: 6.652 min
Delta R.T.: 0.002 min
Response: 68362
Conc: 13.51 ng/ml

Instrument :
ECD_D
ClientSampleId :
TP-28


Manual Integrations
APPROVED

mohammad
4/21/2020 8:26:05 AM




#6 Toxaphene-5

R.T.: 7.516 min
Delta R.T.: -0.004 min
Response: 268413
Conc: 16.45 ng/ml m



#7 Decachlorobiphenyl

R.T.: 7.871 min
Delta R.T.: 0.005 min
Response: 778545
Conc: 1.01 ng/ml



#7 Decachlorobiphenyl

R.T.: 8.897 min
Delta R.T.: -0.002 min
Response: 938390
Conc: 1.05 ng/ml