

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042125\
 Data File : PD088160.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2025 14:06
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 22:43:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.883	98838493	692.4E6	49.484	47.355
28)	SA Decachlor...	9.073	8.075	151.4E6	785.1E6	45.771	42.487
Target Compounds							
2)	A alpha-BHC	4.000	3.395	221.5E6	1088.5E6	51.359	47.612
3)	MA gamma-BHC...	4.331	3.732	211.7E6	1006.1E6	50.566	46.713
4)	MA Heptachlor	4.931	4.086	206.8E6	1005.4E6	51.206	46.986
5)	MB Aldrin	5.272	4.372	198.1E6	982.4E6	50.161	47.222
6)	B beta-BHC	4.516	4.027	80635066	436.8E6	49.671	47.201
7)	B delta-BHC	4.764	4.265	208.6E6	1006.6E6	50.563	47.343
8)	B Heptachlo...	5.692	4.876	175.4E6	881.1E6	49.088	46.632
9)	A Endosulfan I	6.075	5.250	165.6E6	828.3E6	49.047	45.986
10)	B gamma-Chl...	5.947	5.129	177.4E6	942.5E6	48.973	46.439
11)	B alpha-Chl...	6.028	5.194	176.9E6	906.9E6	49.021	46.264
12)	B 4,4'-DDE	6.197	5.378	158.1E6	911.2E6	47.930	46.270
13)	MA Dieldrin	6.348	5.516	176.5E6	922.7E6	49.474	46.292
14)	MA Endrin	6.576	5.792	146.0E6	834.5E6	48.936	45.844
15)	B Endosulfa...	6.787	6.083	148.5E6	798.6E6	47.508	45.579
16)	A 4,4'-DDD	6.706	5.933	122.4E6	751.9E6	48.687	45.887
17)	MA 4,4'-DDT	7.022	6.187	134.8E6	788.4E6	48.520	46.332
18)	B Endrin al...	6.916	6.261	108.3E6	587.3E6	46.919	44.158
19)	B Endosulfa...	7.149	6.485	138.2E6	774.3E6	48.063	45.196
20)	A Methoxychlor	7.493	6.758	71198716	406.6E6	47.480	44.577
21)	B Endrin ke...	7.631	6.994	148.5E6	834.5E6	48.108	44.616
22)	Mirex	8.114	7.188	110.0E6	651.4E6	46.843	43.949

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042125\
Data File : PD088160.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2025 14:06
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 22:43:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

