

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042125\
 Data File : PD088171.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2025 17:19
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 22:45:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.556	2.883	110.6E6	768.6E6	55.377	52.569
28) SA Decachlor...	9.081	8.078	157.1E6	783.4E6	47.489	42.392
Target Compounds						
2) A alpha-BHC	4.006	3.396	252.2E6	1210.6E6	58.491	52.953
3) MA gamma-BHC...	4.337	3.733	239.0E6	1116.9E6	57.082	51.854
4) MA Heptachlor	4.936	4.086	226.3E6	1086.6E6	56.022	50.781
5) MB Aldrin	5.278	4.373	222.2E6	1095.2E6	56.263	52.647
6) B beta-BHC	4.521	4.028	90469694	482.3E6	55.729	52.117
7) B delta-BHC	4.770	4.265	233.3E6	1118.0E6	56.566	52.584
8) B Heptachlo...	5.698	4.877	195.7E6	970.5E6	54.744	51.366
9) A Endosulfan I	6.081	5.251	185.1E6	877.4E6	54.818	48.713
10) B gamma-Chl...	5.953	5.130	201.1E6	1040.9E6	55.508	51.286
11) B alpha-Chl...	6.034	5.195	198.5E6	999.1E6	54.986	50.968
12) B 4,4'-DDE	6.202	5.380	175.2E6	1014.5E6	53.126	51.514
13) MA Dieldrin	6.354	5.517	196.2E6	1011.7E6	54.984	50.757
14) MA Endrin	6.581	5.794	160.1E6	915.4E6	53.670	50.289
15) B Endosulfa...	6.793	6.085	163.1E6	866.3E6	52.156	49.443
16) A 4,4'-DDD	6.712	5.934	140.2E6	827.5E6	55.749	50.497
17) MA 4,4'-DDT	7.028	6.189	139.1E6	817.3E6	50.088	48.029
18) B Endrin al...	6.921	6.263	113.0E6	615.7E6	48.945	46.292
19) B Endosulfa...	7.156	6.487	149.1E6	827.6E6	51.850	48.305
20) A Methoxychlor	7.500	6.759	72975090	418.8E6	48.665	45.914
21) B Endrin ke...	7.636	6.996	157.9E6	878.1E6	51.171	46.947
22) Mirex	8.120	7.190	114.1E6	680.7E6	48.593	45.927

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042125\
Data File : PD088171.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2025 17:19
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 22:45:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

