

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042221\
 Data File : PD062272.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2021 19:40
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 22:09:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032921.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 30 04:47:56 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.484	3.981	38350432	456.4E6	45.803	52.527
2)	SA Decachlor...	8.257	9.042	51675865	402.9E6	41.998	42.681
Target Compounds							
2)	A alpha-BHC	3.924	4.370	59553094	737.9E6	47.152	54.858
3)	MA gamma-BHC...	4.208	4.654	56831045	675.4E6	47.867	54.146
4)	MA Heptachlor	4.511	5.162	59611916	680.5E6	51.206	52.219
5)	MB Aldrin	4.761	5.467	58758255	650.8E6	48.959	52.295
6)	B beta-BHC	4.459	4.824	29492735	280.0E6	54.083	54.205
7)	B delta-BHC	4.665	5.041	58068776	616.8E6	47.801	46.633
8)	B Heptachlo...	5.206	5.852	55893007	546.5E6	48.758	46.157
9)	A Endosulfan I	5.546	6.209	52864588	548.0E6	48.141	50.549
10)	B gamma-Chl...	5.434	6.087	57983728	615.0E6	49.501	51.877
11)	B alpha-Chl...	5.492	6.160	56796532	582.6E6	48.878	51.109
12)	B 4,4'-DDE	5.655	6.316	54353462	575.7E6	48.371	49.957
13)	MA Dieldrin	5.790	6.466	56024165	569.5E6	48.468	49.340
14)	MA Endrin	6.048	6.684	48749300	479.2E6	47.444	48.422
15)	B Endosulfa...	6.322	6.893	48303113	449.9E6	47.130	47.390
16)	A 4,4'-DDD	6.173	6.805	47809293	468.3E6	51.373	49.456
17)	MA 4,4'-DDT	6.414	7.104	45141272	437.5E6	50.116	45.807
18)	B Endrin al...	6.491	7.017	38673539	357.9E6	45.266	45.515
19)	B Endosulfa...	6.705	7.240	48782596	472.3E6	45.209	47.120
20)	A Methoxychlor	6.961	7.563	25097657	207.7E6	53.117	46.070
21)	B Endrin ke...	7.199	7.710	55471812	461.2E6	43.465	45.470
22)	Mirex	7.389	8.166	43330088	294.1E6	51.913	45.643

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042221\
Data File : PD062272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2021 19:40
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 22:09:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032921.M
Quant Title : GC Extractables
QLast Update : Tue Mar 30 04:47:56 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

