

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX042621.M
 Title : GC Extractables
 Last Update : Tue Apr 27 02:33:04 2021
 Response Via : Initial Calibration

Calibration Files

500 =PL066480.D 1000 =PL066478.D 250 =PL066481.D
 750 =PL066479.D 100 =PL066482.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	3.795	3.835	3.752	3.760	4.188	3.866	E6	4.73
2)	Toxaphene-1	2.757	2.633	2.816	2.642	3.233	2.816	E4	8.71
3)	Toxaphene-2	2.832	2.798	2.641	2.753	2.862	2.777	E4	3.10
4)	Toxaphene-3	7.794	7.740	7.772	7.610	8.465	7.876	E4	4.28
5)	Toxaphene-4	5.700	5.920	5.478	5.730	5.454	5.656	E4	3.42
6)	Toxaphene-5	7.850	8.058	7.398	7.615	6.929	7.570	E4	5.75
7) SA	Decachlorobiphenyl	3.798	3.734	4.050	3.792	4.625	4.000	E6	9.25

Signal #2 Calibration Files

500 =PL066480.D 1000 =PL066478.D 250 =PL066481.D
 750 =PL066479.D 100 =PL066482.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	3.575	3.661	3.619	3.661	3.832	3.670	E6	2.65
2)	Toxaphene-1	2.990	2.818	3.041	2.994	3.235	3.015	E4	4.95
3)	Toxaphene-2	3.034	2.953	3.179	3.036	3.577	3.156	E4	7.90
4)	Toxaphene-3	5.630	5.445	5.725	5.571	6.423	5.759	E4	6.69
5)	Toxaphene-4	5.432	5.371	5.673	5.500	5.982	5.592	E4	4.39
6)	Toxaphene-5	5.850	5.541	5.921	5.792	6.322	5.885	E4	4.81
7) SA	Decachlorobiphenyl	3.018	2.715	3.166	2.962	3.525	3.077	E6	9.70

(#) = Out of Range