

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042421\
 Data File : PD062334.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2021 09:24
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 06:48:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042321.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 23 23:19:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.485	3.981	43517339	544.6E6	53.429	51.716
28)	SA Decachlor...	8.256	9.040	58799887	439.5E6	48.474	51.802
Target Compounds							
2)	A alpha-BHC	3.924	4.369	69369333	879.4E6	54.156	52.036
3)	MA gamma-BHC...	4.209	4.653	65357699	803.7E6	53.106	51.170
4)	MA Heptachlor	4.512	5.162	68524975	797.3E6	52.282	50.878
5)	MB Aldrin	4.762	5.467	66867970	762.1E6	53.427	50.696
6)	B beta-BHC	4.459	4.823	30758452	343.1E6	51.090	51.291
7)	B delta-BHC	4.665	5.040	68274604	771.2E6	53.910	55.399
8)	B Heptachlo...	5.206	5.853	64202580	692.6E6	52.293	50.417
9)	A Endosulfan I	5.547	6.209	61235736	652.2E6	51.913	49.801
10)	B gamma-Chl...	5.434	6.086	66270700	712.1E6	52.009	50.449
11)	B alpha-Chl...	5.492	6.159	64944869	675.3E6	51.766	49.191
12)	B 4,4'-DDE	5.656	6.316	63160558	695.4E6	52.032	49.250
13)	MA Dieldrin	5.791	6.466	66867744	699.8E6	52.399	49.756
14)	MA Endrin	6.049	6.684	56931740	563.1E6	49.124	46.159
15)	B Endosulfa...	6.323	6.893	54638535	576.0E6	48.354	50.057
16)	A 4,4'-DDD	6.173	6.805	51893143	560.0E6	51.854	50.254
17)	MA 4,4'-DDT	6.415	7.104	50271172	547.2E6	49.229	54.877
18)	B Endrin al...	6.491	7.016	48629143	494.7E6	49.223	52.407
19)	B Endosulfa...	6.705	7.240	58366519	596.4E6	49.433	54.979
20)	A Methoxychlor	6.961	7.563	30768255	264.8E6	51.779	54.203
21)	B Endrin ke...	7.199	7.710	67935224	543.7E6	55.770	54.193
22)	Mirex	7.389	8.165	53479999	281.7E6	53.486	45.069

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042421\
Data File : PD062334.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2021 09:24
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 06:48:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042321.M
Quant Title : GC Extractables
QLast Update : Fri Apr 23 23:19:57 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

