

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042421\
 Data File : PD062385.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2021 21:47
 Operator : AR\AJ
 Sample : PB135947BS
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB135947BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 06:59:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042321.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 23 23:19:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.485	3.981	15660371	206.0E6	19.227	19.556
28)	SA Decachlor...	8.257	9.042	23425156	179.2E6	19.311	21.120
Target Compounds							
2)	A alpha-BHC	3.924	4.370	59425381	788.7E6	46.393	46.665
3)	MA gamma-BHC...	4.209	4.655	56764515	718.5E6	46.124	45.746
4)	MA Heptachlor	4.512	5.163	60151353	734.1E6	45.893	46.844
5)	MB Aldrin	4.762	5.468	59214007	703.5E6	47.312	46.803
6)	B beta-BHC	4.459	4.824	25274615	301.5E6	41.981	45.064
7)	B delta-BHC	4.665	5.041	58175247	695.2E6	45.935	49.399
8)	B Heptachlo...	5.206	5.853	56774398	633.2E6	46.243	46.093
9)	A Endosulfan I	5.547	6.210	53597833	594.1E6	45.438	45.362
10)	B gamma-Chl...	5.434	6.087	58350719	647.0E6	45.793	45.832
11)	B alpha-Chl...	5.492	6.160	57682203	618.4E6	45.977	45.048
12)	B 4,4'-DDE	5.655	6.316	55235848	623.0E6	45.503	44.124
13)	MA Dieldrin	5.791	6.467	56944021	622.9E6	44.623	44.287
14)	MA Endrin	6.049	6.685	50262027	494.9E6	43.369	40.568
15)	B Endosulfa...	6.322	6.893	46572476	497.2E6	41.215	43.207
16)	A 4,4'-DDD	6.173	6.806	47616765	508.0E6	47.581	45.584
17)	MA 4,4'-DDT	6.415	7.104	44956937	494.1E6	44.025	49.547
18)	B Endrin al...	6.491	7.017	39819721	404.2E6	40.306	42.818
19)	B Endosulfa...	6.705	7.240	47405913	493.2E6	40.150	45.461
20)	A Methoxychlor	6.961	7.563	25577606	227.8E6	43.043	46.617
21)	B Endrin ke...	7.199	7.711	56684812	486.1E6	46.534	48.455
22)	Mirex	7.390	8.166	45235216	308.3E6	45.240	49.327

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042421\
Data File : PD062385.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2021 21:47
Operator : AR\AJ
Sample : PB135947BS
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB135947BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 06:59:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042321.M
Quant Title : GC Extractables
QLast Update : Fri Apr 23 23:19:57 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

