

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042523\
 Data File : PD074907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2023 20:55
 Operator : AR\AJ
 Sample : 02418-07
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 EW8Z3

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/28/2023
 Supervised By : Ankita Jodhani 05/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 25 22:20:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041423CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 14 02:10:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.769	31403681	17795722	12.301	12.650
27)	SA Decachlor...	9.091	7.919	93106726	39531645	30.374	29.290
Target Compounds							
4)	MA Heptachlor	4.930	3.943	78853044	37370275	19.308	19.672
5)	MB Aldrin	5.272	4.224	92074244	44192474	23.391m	23.356
8)	B Heptachlo...	5.700	4.727	112.3E6	45808276	30.698	26.329
10)	B trans-Chl...	5.957	4.976	507.5E6	218.4E6	143.912	127.931m
11)	B cis-Chlor...	6.039	5.041	222.5E6	68651534	60.573	39.718 #
12)	B 4,4'-DDE	6.206	5.231	37644069	16603100	10.899m	10.169
13)	MA Dieldrin	6.363	5.362	797.1E6	345.8E6	213.252	194.646
16)	A 4,4'-DDD	6.725	5.786	55785124	32890735	19.586m	23.781
17)	MA 4,4'-DDT	7.042	6.036	52628680	26865052	17.295	18.683

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042523\
Data File : PD074907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2023 20:55
Operator : AR\AJ
Sample : 02418-07
Misc :
ALS Vial : 18 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 22:20:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041423CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 14 02:10:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
EW823
Manual Integrations
APPROVED
Reviewed By :Abdul Mirza 04/28/2023
Supervised By :Ankita Jodhani 05/01/2023