

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042523\
 Data File : PD074915.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2023 22:42
 Operator : AR\AJ
 Sample : 02418-25
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 EW8Z9

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/01/2023
 Supervised By : Ankita Jodhani 05/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 28 07:33:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041423CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 14 02:10:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.769	26506572	16109625	10.382	11.451
27)	SA Decachlor...	9.079	7.918	28775207	28190762	9.387	20.887 #
Target Compounds							
8)	B Heptachlo...	5.695	4.725	13111435	5189223	3.583m	2.983
9)	A Endosulfan I	6.118	5.121	114.5E6	74015596	33.836m	47.497m#
10)	B trans-Chl...	5.955	4.975	113.9E6	47287924	32.299m	27.695m
11)	B cis-Chlor...	6.037	5.041	114.4E6	33750058	31.137	19.526 #
12)	B 4,4'-DDE	6.205	5.232	21783782	17636663	6.307m	10.802 #
13)	MA Dieldrin	6.363	5.365	141.0E6	93777490	37.724	52.793 #
16)	A 4,4'-DDD	6.716	5.784	70648344	17831876	24.805m	12.893 #
17)	MA 4,4'-DDT	7.040	6.033	25354907	15017208	8.332m	10.443m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042523\
Data File : PD074915.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2023 22:42
Operator : AR\AJ
Sample : 02418-25
Misc :
ALS Vial : 26 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 07:33:25 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041423CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 14 02:10:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

