

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042621\
 Data File : PD062433.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2021 19:40
 Operator : AR\AJ
 Sample : PB136002BS
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB136002BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 01:36:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042321.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 23 23:19:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.485	3.981	18820098	258.9E6	23.107	24.588
28)	SA Decachlor...	8.256	9.040	27773409	212.2E6	22.896	25.007
Target Compounds							
2)	A alpha-BHC	3.924	4.370	75295017	1005.3E6	58.783	59.484
3)	MA gamma-BHC...	4.209	4.654	71392003	921.6E6	58.010	58.677
4)	MA Heptachlor	4.511	5.162	74998450	916.5E6	57.221	58.487
5)	MB Aldrin	4.761	5.467	73989645	880.0E6	59.117	58.541
6)	B beta-BHC	4.459	4.824	31621092	382.7E6	52.523	57.198
7)	B delta-BHC	4.665	5.040	73736453	912.4E6	58.222	66.549
8)	B Heptachlo...	5.205	5.853	70396598	790.7E6	57.338	57.560
9)	A Endosulfan I	5.545	6.209	66540952	739.9E6	56.410	56.496
10)	B gamma-Chl...	5.433	6.086	71883962	804.1E6	56.414	56.963
11)	B alpha-Chl...	5.492	6.160	71078133	776.9E6	56.655	56.590
12)	B 4,4'-DDE	5.654	6.316	68083354	785.3E6	56.087	55.618
13)	MA Dieldrin	5.790	6.466	69947669	779.7E6	54.813	55.432
14)	MA Endrin	6.049	6.683	63030327	630.7E6	54.386	51.697
15)	B Endosulfa...	6.322	6.892	59899992	605.8E6	53.010	52.642
16)	A 4,4'-DDD	6.172	6.806	59350785	622.7E6	59.306	55.876
17)	MA 4,4'-DDT	6.414	7.105	57736906	617.5E6	56.540	61.925
18)	B Endrin al...	6.490	7.016	49071417	499.1E6	49.670	52.879
19)	B Endosulfa...	6.704	7.240	60806913	621.7E6	51.499	57.305
20)	A Methoxychlor	6.961	7.562	31936986	281.5E6	53.745	57.627
21)	B Endrin ke...	7.198	7.710	70051300	608.5E6	57.507	60.654
22)	Mirex	7.389	8.165	54534586	382.7E6	54.540	61.224

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042621\
Data File : PD062433.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2021 19:40
Operator : AR\AJ
Sample : PB136002BS
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB136002BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 01:36:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042321.M
Quant Title : GC Extractables
QLast Update : Fri Apr 23 23:19:57 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

