

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042622\
 Data File : PD069241.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2022 16:53
 Operator : AR\AJ
 Sample : INDB355
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB3184

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 21:56:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041422CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 21 02:24:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.450	4.172	24674646	233.1E6	19.475	19.317
27)	SA Decachlor...	8.207	9.370	52354067	273.1E6	39.370	38.833
Target Compounds							
5)	MB Aldrin	4.718	5.699	33495584	293.4E6	21.027	20.555
6)	B beta-BHC	4.424	5.022	15873982	135.8E6	20.494	20.246
7)	B delta-BHC	4.628	5.248	33492840	282.7E6	20.983	24.078
8)	B Heptachlo...	5.162	6.088	31154653	252.7E6	21.164	22.684
10)	B trans-Chl...	5.388	6.324	31936378	261.5E6	21.303	20.609
11)	B cis-Chlor...	5.447	6.399	31891035	257.2E6	21.365	20.654
12)	B 4,4'-DDE	5.613	6.549	58195603	450.8E6	43.387	41.434
15)	B Endosulfa...	6.277	7.135	55330259	393.2E6	42.937	41.026
18)	B Endrin al...	6.446	7.258	42882221	296.3E6	43.201	41.465
19)	B Endosulfa...	6.659	7.483	54626723	365.0E6	43.597	41.602
21)	B Endrin ke...	7.152	7.959	62578594	381.7E6	43.316	41.345

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042622\
Data File : PD069241.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2022 16:53
Operator : AR\AJ
Sample : INDB355
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 21:56:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041422CLP.M
Quant Title : GC Extractables
QLast Update : Thu Apr 21 02:24:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

