

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042623\
 Data File : PD074927.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2023 11:47
 Operator : AR\AJ
 Sample : 02418-02DL 2X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 EW8Y8DL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/27/2023
 Supervised By : Ankita Jodhani 05/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 04:50:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041423CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 14 02:10:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.767	11737451	7528526	4.597m	5.352
27)	SA Decachlor...	9.075	7.916	80217394	19686539	26.169m	14.586 #
Target Compounds							
8)	B Heptachlo...	5.697	4.725	29023042	6800167	7.932m	3.908 #
10)	B trans-Chl...	5.958	4.974	72259859	48219187	20.490m	28.240m#
11)	B cis-Chlor...	6.040	5.040	65879876	29872720	17.934	17.283
12)	B 4,4'-DDE	6.209	5.231	30160285	15506818	8.732m	9.497
13)	MA Dieldrin	6.366	5.362	293.2E6	164.6E6	78.437	92.661
14)	MA Endrin	6.595	5.637	46431378	28967681	14.415	18.455 #
16)	A 4,4'-DDD	6.727	5.782	42662913	10719050	14.979	7.750m#
17)	MA 4,4'-DDT	7.042	6.034	43184346	34841171	14.191m	24.230 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042623\
Data File : PD074927.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2023 11:47
Operator : AR\AJ
Sample : 02418-02DL 2X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 04:50:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041423CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 14 02:10:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integrations
APPROVED
Reviewed By: Abdul Mirza 04/27/2023
Supervised By: Ankita Jodhani 05/01/2023

Instrument :
ECD_D
Client Sampled :
EW8Y8DL