

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042723\
 Data File : PD075005.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2023 03:36
 Operator : AR\AJ
 Sample : 02418-22
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 EW8Z6

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/28/2023
 Supervised By : Ankita Jodhani 05/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 28 06:00:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042723CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 28 02:34:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	2.771	32511358	15562444	14.554m	15.012
27)	SA Decachlor...	9.084	7.917	48609073	27473688	20.352m	28.600 #
Target Compounds							
9)	A Endosulfan I	6.116	5.122	19644123	18243241	6.746m	15.989m#
10)	B trans-Chl...	5.953	4.977	19710010	9057426	6.584m	7.477m
11)	B cis-Chlor...	6.035	5.039	28621705	7758296	9.182	6.266m#
12)	B 4,4'-DDE	6.207	5.231	200.6E6	91028017	68.309	80.472
16)	A 4,4'-DDD	6.722	5.784	67608848	22978031	27.958m	23.782m
17)	MA 4,4'-DDT	7.037	6.036	46728543	21995517	18.598m	22.173

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042723\
Data File : PD075005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2023 03:36
Operator : AR\AJ
Sample : 02418-22
Misc :
ALS Vial : 41 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 06:00:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042723CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 28 02:34:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integrations
APPROVED
Reviewed By: Abul Mirza 04/28/2023
Supervised By: Ankita Jodhani 05/01/2023

Instrument :
ECD_D
Client/Sample :
EW826