Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042823\
Data File : PD075031.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2023 15:31
Operator : AR\AJ
Sample : 02417-14
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
EW8Y0

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 05/02/2023
Supervised By :Ankita Jodhani 05/02/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 22:16:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042723CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 28 02:34:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	3.548	2.771	40231891	18157247	18.011	17.515
27) SA	Decachlor...	9.082	7.917	121.4E6	30675247	50.827m	31.933 #
	Target Compounds						
4) MA	Heptachlor	4.927	3.943	5446399	6538260	1.572m	4.688m#
5) MB	Aldrin	5.267	4.238	3343928	5171657	0.998m	3.933m#
7) B	delta-BHC	4.787	4.138	5113758	2032438	1.708m	1.445m
8) B	Heptachlo...	5.691	4.725	34317532	3605075	11.427m	2.936m#
10) B	trans-Chl...	5.954	4.976	69614775	20522033	23.256m	16.940m#
11) B	cis-Chlor...	6.037	5.039	63165829	18327675	20.264	14.802m#
12) B	4,4'-DDE	6.204	5.230	29226418	8003333	9.950m	7.075m#
13) MA	Dieldrin	6.361	5.363	103.6E6	69237211	32.606	55.159 #
14) MA	Endrin	6.589	5.645	35161407	5339231	12.979	4.831 #
16) A	4,4'-DDD	6.733	5.781	25389065	2805927	10.499m	2.904m#
17) MA	4,4'-DDT	7.039	6.032	30654726	32067183	12.200m	32.326m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042823\
Data File : PD075031.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2023 15:31
Operator : AR\AJ
Sample : 02417-14
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 22:16:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042723CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 28 02:34:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





Manual Integrations APPROVED
Reviewed By :Abdul Mirza 05/02/2023
Supervised By :Ankita Jodhani 05/02/2023

Instrument : ECD_D
ClientSampleId : EW8Y0