

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042823\
 Data File : PD075030.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2023 15:04
 Operator : AR\AJ
 Sample : 02417-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 EW8X4

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/01/2023
 Supervised By : Ankita Jodhani 05/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 28 22:16:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042723CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 28 02:34:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	2.771	27712169	14471954	12.406m	13.960
2)	SA Decachlor...	9.097	7.917	67750272	27395806	28.366m	28.519m
Target Compounds							
4)	MA Heptachlor	4.930	3.945	15446.8E6	4554.7E6	4458.927	3265.615 #
5)	MB Aldrin	5.271	4.225	8701.1E6	2670.1E6	2597.981m	2030.561
6)	B beta-BHC	4.532	3.898	17159662	7285889	11.641	10.706
8)	B Heptachlo...	5.699	4.727	6583.7E6	1961.2E6	2192.225	1597.312 #
10)	B trans-Chl...	5.956	4.978	17076.7E6	5573.2E6	5704.704	4600.450
11)	B cis-Chlor...	6.037	5.042	5746.8E6	1371.7E6	1843.635	1107.871 #
12)	B 4,4'-DDE	6.203	5.229	644.6E6	206.0E6	219.454m	182.084m
13)	MA Dieldrin	6.363	5.363	28387.0E6	10719.9E6	8930.965	8540.213
14)	MA Endrin	6.592	5.646	210.5E6	201.7E6	77.695m	182.476 #
16)	A 4,4'-DDD	6.723	5.784	396.9E6	176.0E6	164.150m	182.119m
17)	MA 4,4'-DDT	7.040	6.037	1786.7E6	619.8E6	711.080	624.846
21)	B Endrin ke...	7.662	6.841	1956.4E6	665.5E6	676.237	558.115

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042823\
Data File : PD075030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2023 15:04
Operator : AR\AJ
Sample : 02417-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 22:16:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042723CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 28 02:34:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
EW8X4

Manual Integrations
APPROVED

Reviewed By: Abdul Miza 05/01/2023
Supervised By: Ankita Jodhani 05/02/2023