

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042823\
 Data File : PD075036.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2023 17:32
 Operator : AR\AJ
 Sample : INDA399
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3199

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 05/01/2023
 Supervised By :Ankita Jodhani 05/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 28 22:17:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042723CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 28 02:34:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.770	47727927	22290627	21.366	21.502
2)	SA Decachlor...	9.090	7.918	78956415	35169732	33.058	36.612
Target Compounds							
2)	A alpha-BHC	4.004	3.273	79898686	31204432	21.771	21.670
3)	MA gamma-BHC...	4.337	3.603	73606361	30292109	21.528	21.769
4)	MA Heptachlor	4.929	3.944	74658235	29200738	21.551	20.936
9)	A Endosulfan I	6.085	5.097	60380719	23636393	20.735m	20.716
13)	MA Dieldrin	6.360	5.361	134.6E6	52647972	42.356m	41.943m
14)	MA Endrin	6.592	5.637	113.1E6	44595050	41.747	40.349m
16)	A 4,4'-DDD	6.725	5.786	104.3E6	41643738	43.121	43.102
17)	MA 4,4'-DDT	7.040	6.037	97344281	38547711	38.743	38.859
20)	A Methoxychlor	7.516	6.613	279.2E6	106.9E6	193.646	193.470

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042823\
Data File : PD075036.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2023 17:32
Operator : AR\AJ
Sample : INDA399
Misc :
ALS Vial : 40 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 22:17:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042723CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 28 02:34:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

