

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042920\
 Data File : PD057843.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Apr 2020 05:58
 Operator : AJ\MA
 Sample : INDA501
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 07:18:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042920CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 30 06:57:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.228	3.896	62879838	102.4E6	74.301	73.405
27)	SA Decachlor...	7.872	8.900	132.4E6	190.0E6	138.017	131.720
Target Compounds							
2)	A alpha-BHC	3.651	4.279	100.9E6	160.6E6	85.833	83.201
3)	MA gamma-BHC...	3.923	4.560	93699282	149.5E6	82.099	81.339
4)	MA Heptachlor	4.203	5.062	85698085	149.4E6	81.353	78.001
9)	A Endosulfan I	5.200	6.100	71857796	122.1E6	79.734	76.784
13)	MA Dieldrin	5.438	6.354	170.9E6	276.9E6	164.851	159.195
14)	MA Endrin	5.690	6.571	145.7E6	234.7E6	158.722	144.992
16)	A 4,4'-DDD	5.827	6.697	134.7E6	218.0E6	164.186	159.275
17)	MA 4,4'-DDT	6.061	6.995	139.9E6	231.0E6	167.927	159.211
20)	A Methoxychlor	6.609	7.453	325.5E6	555.0E6	739.037	687.856

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042920\
Data File : PD057843.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Apr 2020 05:58
Operator : AJ\MA
Sample : INDA501
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 07:18:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042920CLP.M
Quant Title : GC Extractables
QLast Update : Thu Apr 30 06:57:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

